

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX091620.M
 Title : GC Extractables
 Last Update : Wed Sep 16 04:05:05 2020
 Response Via : Initial Calibration

Calibration Files

500 =PD059227.D 1000 =PD059225.D 750 =PD059226.D
 250 =PD059228.D 100 =PD059229.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	9.003	9.097	9.052	8.874	8.753	8.956	E5	1.57
2)	Toxaphene-1	1.016	0.968	1.008	1.013	0.996	1.000	E4	1.97
3)	Toxaphene-2	1.546	1.555	1.568	1.564	1.659	1.578	E4	2.92
4)	Toxaphene-3	1.588	1.541	1.570	1.541	1.631	1.574	E4	2.38
5)	Toxaphene-4	7.193	7.371	7.663	6.777	4.448	6.691	E3	19.34
6)	Toxaphene-5	6.024	5.996	6.348	5.346	5.426	5.828	E3	7.33
7) SA	Decachlorobiphenyl	9.277	9.039	9.245	9.394	9.696	9.330	E5	2.59

Signal #2 Calibration Files

500 =PD059227.D 1000 =PD059225.D 750 =PD059226.D
 250 =PD059228.D 100 =PD059229.D

	Compound	500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.694	1.715	1.697	1.705	1.684	1.699	E6	0.69
2)	Toxaphene-1	1.855	1.906	1.883	1.814	1.789	1.849	E4	2.60
3)	Toxaphene-2	2.154	1.922	2.034	2.139	2.316	2.113	E4	6.96
4)	Toxaphene-3	4.447	4.460	4.551	4.294	4.215	4.393	E4	3.09
5)	Toxaphene-4	2.844	2.862	2.898	2.653	2.601	2.772	E4	4.85
6)	Toxaphene-5	2.271	2.097	2.178	2.213	1.968	2.145	E4	5.47
7) SA	Decachlorobiphenyl	1.550	1.483	1.517	1.599	1.801	1.590	E6	7.90

(#) = Out of Range