

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091719\
 Data File : PD054647.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Sep 2019 11:37
 Operator : SG\AJ
 Sample : K4810-19
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AL9

Manual Integrations
 APPROVED

Ankita
 9/18/2019 3:26:05 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 18 03:48:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091619CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 17 05:19:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.295	3.967	6115715	7088553	6.730	7.316
27) SA Decachlor...	7.974	9.000	9966251	17022429	9.134	15.198m#
Target Compounds						
4) MA Heptachlor	4.272	5.139	1505493	767687	1.277m	0.571m#
15) B Endosulfa...	6.065	6.853	4276757	2457585	3.562m	1.959m#
16) A 4,4'-DDD	5.922	6.774	4892865	8244402	4.508m	7.084 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091719\
Data File : PD054647.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Sep 2019 11:37
Operator : SG\AJ
Sample : K4810-19
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
C0AL9

Manual Integrations
APPROVED

Ankita
9/18/2019 3:26:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 03:48:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091619CLP.M
Quant Title : GC Extractables
QLast Update : Tue Sep 17 05:19:07 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

