

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091720\
 Data File : PD059253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2020 14:10
 Operator : DD\AJ
 Sample : PB131623BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB131623BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 01:25:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX091620.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 16 04:05:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.205	3.865	17618600	33484678	19.673	19.710
7) SA Decachlor...	7.838	8.858	19646451	34122727	21.057	21.459


Target Compounds


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091720\
Data File : PD059253.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2020 14:10
Operator : DD\AJ
Sample : PB131623BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 01:25:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX091620.M
Quant Title : GC Extractables
QLast Update : Wed Sep 16 04:05:05 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.205 min
Delta R.T.: 0.000 min
Response: 17618600
Conc: 19.67 ng/ml

Instrument :
ECD_D
ClientSampleId :
PB131623BL




#1 Tetrachloro-m-xylene

R.T.: 3.865 min
Delta R.T.: 0.000 min
Response: 33484678
Conc: 19.71 ng/ml



#7 Decachlorobiphenyl

R.T.: 7.838 min
Delta R.T.: 0.000 min
Response: 19646451
Conc: 21.06 ng/ml



#7 Decachlorobiphenyl

R.T.: 8.858 min
Delta R.T.: 0.000 min
Response: 34122727
Conc: 21.46 ng/ml