

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.535min 9.172 ng/ml

response 9276949

(1) Tetrachloro-m-xylene #2 (SA)

2.902min 18.135 ng/ml

response 64176745

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.533min 20.395 ng/ml m

response 20627726

(1) Tetrachloro-m-xylene #2 (SA)

2.901min 17.919 ng/ml m

response 63410419

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)

6.072min 1.041 ng/ml

response 1432401

(9) Endosulfan I #2 (A)

5.277min 2.854 ng/ml

response 10941816

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

DDC07RE


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024

Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)

6.063min 1.920 ng/ml m

response 2641458

(9) Endosulfan I #2 (A)

5.276min 2.480 ng/ml m

response 9507457

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(10) trans-Chlordane (B)

5.913min 2.536 ng/ml

response 3576756

(10) trans-Chlordane #2 (B)

5.144min 5.435 ng/ml

response 22354812

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(10) trans-Chlordane (B)

5.912min 3.366 ng/ml m

response 4746662

(10) trans-Chlordane #2 (B)

5.142min 4.605 ng/ml m

response 18941435

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) cis-Chlordane (B)

6.072min 0.969 ng/ml

response 1432401

(11) cis-Chlordane #2 (B)

5.219min 8.049 ng/ml

response 33770869

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) cis-Chlordane (B)
5.995min 2.127 ng/ml m
response 3145145

(11) cis-Chlordane #2 (B)
5.217min 5.008 ng/ml m
response 21010451

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.035min 10.313 ng/ml

response 13468533

(27) Decachlorobiphenyl #2 (SA)

8.101min 9.369 ng/ml

response 32577435

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 19:50
Operator : AR\AJ
Sample : P3910-05RE
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DDC07RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/19/2024
Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 18 22:34:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

9.033min 12.756 ng/ml m

response 16659698

(27) Decachlorobiphenyl #2 (SA)

8.100min 12.471 ng/ml m

response 43365785