

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
 Data File : PD085349.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2024 23:24
 Operator : AR\AJ
 Sample : P3910-21MSRE
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDC66MSRE

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/19/2024
 Supervised By :Ankita Jodhani 09/19/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 02:56:23 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 27 02:47:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.534	2.912	14910937	324.2E6	14.743m	91.604m#
2)	SA Decachlor...	9.033	8.100	32403671	45005876	24.811m	12.943m#
Target Compounds							
2)	A alpha-BHC	3.982	3.412	94240253	253.8E6	55.430	50.049
3)	MA gamma-BHC...	4.308	3.747	175.1E6	233.8E6	108.255m	48.251 #
4)	MA Heptachlor	4.909	4.086	74711054	758.3E6	48.558	178.120m#
5)	MB Aldrin	5.250	4.388	77563851	206.8E6	48.497	46.553
6)	B beta-BHC	4.495	4.041	30421611	90310383	41.882	41.657
7)	B delta-BHC	4.743	4.279	49974076	260.7E6	30.472m	52.508 #
8)	B Heptachlo...	5.669	4.891	69013870	192.1E6	46.272	45.358
9)	A Endosulfan I	6.050	5.267	59087800	174.4E6	42.948m	45.492
10)	B trans-Chl...	5.923	5.143	65360493	189.9E6	46.343	46.173m
11)	B cis-Chlor...	6.003	5.210	65807832	190.7E6	44.508m	45.449
12)	B 4,4'-DDE	6.173	5.395	122.0E6	368.7E6	97.808	93.982
13)	MA Dieldrin	6.324	5.534	137.9E6	392.5E6	93.480	91.354
14)	MA Endrin	6.551	5.812	117.4E6	357.9E6	104.025	96.659
15)	B Endosulfa...	6.763	6.104	111.4E6	324.4E6	84.406	87.503
16)	A 4,4'-DDD	6.681	5.952	100.7E6	320.3E6	102.012	94.378
17)	MA 4,4'-DDT	6.997	6.207	84877659	276.9E6	94.822	86.558
18)	B Endrin al...	6.891	6.283	84450642	239.7E6	93.511	83.935
19)	B Endosulfa...	7.125	6.508	106.3E6	307.9E6	92.299	86.223
20)	A Methoxychlor	7.468	6.779	259.8E6	657.3E6	498.750	397.351
21)	B Endrin ke...	7.604	7.019	116.1E6	360.7E6	94.971	81.926

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085349.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2024 23:24
Operator : AR\AJ
Sample : P3910-21MSRE
Misc :
ALS Vial : 32 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 02:56:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Miza 09/19/2024
Supervised By: Ankita Jodhani 09/19/2024

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
DDC66MSRE