

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
 Data File : PD085365.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2024 02:57
 Operator : AR\AJ
 Sample : P4016-15
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDC20

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/20/2024
 Supervised By :Ankita Jodhani 09/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 05:19:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 27 02:47:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.529	2.913	41537869	6378.4E6	41.070m	1802.424 #
27)	SA Decachlor...	9.033	8.100	23457501	41464051	17.961m	11.925 #
Target Compounds							
2)	A alpha-BHC	3.980	3.409	15998107	74397926	9.410m	14.673m#
7)	B delta-BHC	4.760	4.276	605.2E6	94184574	369.052m	18.973m#
9)	A Endosulfan I	6.052	5.274	10374353	43644057	7.541m	11.386 #
10)	B trans-Chl...	5.901	5.136	10883982	41530210	7.717m	10.097m#
11)	B cis-Chlor...	5.994	5.218	17962346	95993883	12.149m	22.880 #
12)	B 4,4'-DDE	6.174	5.392	35351217	41282469	28.339	10.524m#
13)	MA Dieldrin	6.326	5.539	8978716	37061919	6.087m	8.627m#
16)	A 4,4'-DDD	6.682	5.957	11135222	33510093	11.278m	9.875m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091824\
Data File : PD085365.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2024 02:57
Operator : AR\AJ
Sample : P4016-15
Misc :
ALS Vial : 48 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 05:19:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Mirza 09/20/2024
Supervised By: Ankita Jodhani 09/20/2024

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
DDC20