

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD091918CLP.M
 Title : GC Extractables
 Last Update : Thu Sep 20 03:57:09 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049291.D 1 =PD049287.D 2 =PD049289.D
 4 =PD049293.D 5 =PD049295.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.285	1.307	1.308	1.259	1.246	1.281	E6	2.17
2) A	alpha-BHC	2.032	2.003	2.027	2.041	2.071	2.035	E6	1.21
3) MA	gamma-BHC (Lindane)	1.926	2.013	1.954	1.904	1.907	1.941	E6	2.32
4) MA	Heptachlor	1.955	1.922	1.969	1.924	1.916	1.937	E6	1.22
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	1.699	1.695	1.725	1.649	1.617	1.677	E6	2.58
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	1.927	1.907	1.922	1.904	1.893	1.911	E6	0.71
14) MA	Endrin	1.710	1.741	1.717	1.681	1.651	1.700	E6	2.04
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD	1.565	1.567	1.569	1.543	1.537	1.556	E6	0.99
17) MA	4,4'-DDT	1.640	1.586	1.618	1.630	1.624	1.620	E6	1.26
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	8.703	8.995	8.956	8.382	7.962	8.600	E5	5.03
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00
24)	Toxaphene-3						0.000		-1.00
25)	Toxaphene-4						0.000		-1.00
26)	Toxaphene-5						0.000		-1.00
27) SA	Decachlorobiphenyl	1.575	1.658	1.638	1.527	1.473	1.574	E6	4.88

Signal #2 Calibration Files

3 =PD049291.D 1 =PD049287.D 2 =PD049289.D
 4 =PD049293.D 5 =PD049295.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.992	2.098	2.048	1.866	1.753	1.952	E7	7.21
2) A	alpha-BHC	3.091	3.269	3.206	2.943	2.791	3.060	E7	6.37
3) MA	gamma-BHC (Lindane)	2.850	3.206	3.036	2.672	2.521	2.857	E7	9.59
4) MA	Heptachlor	2.645	2.691	2.732	2.499	2.354	2.584	E7	6.03
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	2.324	2.507	2.448	2.138	1.988	2.281	E7	9.48
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	2.379	2.598	2.524	2.229	2.074	2.361	E7	9.06
14) MA	Endrin	1.931	2.096	2.042	1.804	1.672	1.909	E7	9.08
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD	1.399	1.447	1.458	1.327	1.304	1.387	E7	5.02
17) MA	4,4'-DDT	1.467	1.467	1.511	1.416	1.354	1.443	E7	4.17
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	5.896	6.045	6.313	5.497	4.967	5.743	E6	9.14

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD091918CLP.M
Title : GC Extractables
Last Update : Thu Sep 20 03:57:09 2018
Response Via : Initial Calibration

Calibration Files

3	=PD049291.D	1	=PD049287.D	2	=PD049289.D
4	=PD049293.D	5	=PD049295.D		

	Compound	3	1	2	4	5	Avg	%RSD
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.458	1.653	1.577	1.363	1.242	1.459	E7 11.27

(#) = Out of Range