

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092023\
 Data File : PD078133.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2023 23:11
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 03:57:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090323.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 01 16:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.759	134.9E6	87026523	53.138	55.023
2)	SA Decachlor...	9.080	7.896	149.3E6	80321835	49.020	46.546
Target Compounds							
2)	A alpha-BHC	3.998	3.260	226.6E6	129.2E6	55.070	55.825
3)	MA gamma-BHC...	4.331	3.590	215.6E6	123.3E6	54.497	55.150
4)	MA Heptachlor	4.921	3.929	204.7E6	121.3E6	52.243	53.938
5)	MB Aldrin	5.264	4.208	206.9E6	123.3E6	51.434	54.116
6)	B beta-BHC	4.529	3.887	84491813	49787546	52.356	55.861
7)	B delta-BHC	4.777	4.117	210.9E6	124.1E6	57.215	57.251
8)	B Heptachlo...	5.691	4.710	176.3E6	110.2E6	49.302	54.315
9)	A Endosulfan I	6.079	5.080	158.7E6	100.6E6	48.250	53.712
10)	B gamma-Chl...	5.949	4.960	171.4E6	108.6E6	48.697	54.019
11)	B alpha-Chl...	6.028	5.024	170.9E6	107.4E6	48.036	53.318
12)	B 4,4'-DDE	6.200	5.213	162.9E6	102.5E6	50.590	56.153
13)	MA Dieldrin	6.354	5.345	170.5E6	107.6E6	47.912	52.710
14)	MA Endrin	6.585	5.621	141.1E6	93905551	46.526	51.844
15)	B Endosulfa...	6.805	5.915	137.5E6	88724750	44.372	51.854
16)	A 4,4'-DDD	6.719	5.768	125.1E6	81219205	48.817	58.121
17)	MA 4,4'-DDT	7.035	6.019	116.6E6	72556219	44.602	51.355
18)	B Endrin al...	6.936	6.095	102.3E6	65880055	43.067	48.806
19)	B Endosulfa...	7.171	6.317	123.9E6	80722319	43.661	50.413
20)	A Methoxychlor	7.510	6.594	63922009	39114199	41.327	50.520
21)	B Endrin ke...	7.658	6.824	145.5E6	94607510	44.763	50.037
22)	Mirex	8.132	7.005	105.6E6	72247486	43.890	47.181

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092023\
Data File : PD078133.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2023 23:11
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 03:57:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090323.M
Quant Title : GC Extractables
QLast Update : Fri Sep 01 16:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

