

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092024\
 Data File : PD085474.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2024 02:11
 Operator : AR\AJ
 Sample : P3927-09
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDCC6

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/23/2024
 Supervised By :Ankita Jodhani 09/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 06:12:58 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 20 05:48:22 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.486	2.904	33107739	1740.7E6	27.340m	460.520m#
2)	SA Decachlor...	8.958	8.087	20411390	48195939	12.443m	18.029m#
Target Compounds							
2)	A alpha-BHC	3.934	3.404	118.7E6	268.6E6	50.229	48.640
4)	MA Heptachlor	4.857	4.089	114.5E6	258.1E6	54.217	54.368m
5)	MB Aldrin	5.195	4.378	116.1E6	254.2E6	53.472	50.141
7)	B delta-BHC	4.692	4.269	149.4E6	235.0E6	65.558	44.612 #
8)	B Heptachlo...	5.612	4.879	536.7E6	313.4E6	268.668m	67.513m#
9)	A Endosulfan I	5.993	5.257	160.5E6	156.1E6	90.541m	37.992 #
11)	B cis-Chlor...	5.945	5.200	157.1E6	205.7E6	78.350m	44.345 #
12)	B 4,4'-DDE	6.119	5.384	190.2E6	380.8E6	108.777	88.873
13)	MA Dieldrin	6.268	5.523	234.7E6	364.1E6	121.404	78.216 #
14)	MA Endrin	6.494	5.801	191.7E6	455.8E6	119.262	109.896
15)	B Endosulfa...	6.706	6.093	148.5E6	347.6E6	81.009	93.943
16)	A 4,4'-DDD	6.627	5.941	155.6E6	305.3E6	118.089	91.164
19)	B Endosulfa...	7.066	6.495	127.7E6	331.0E6	84.552m	96.375m
20)	A Methoxychlor	7.412	6.767	566.7E6	737.3E6	774.510m	454.324 #
21)	B Endrin ke...	7.546	7.007	99831782	366.6E6	62.445m	78.919 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092024\
Data File : PD085474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2024 02:11
Operator : AR\AJ
Sample : P3927-09
Misc :
ALS Vial : 42 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 06:12:58 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

