

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092123\
 Data File : PD078144.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2023 13:45
 Operator : AR\AJ
 Sample : 04450-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 00:22:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 06 16:40:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.759	53881586	33461235	21.232	22.641
27)	SA Decachlor...	9.082	7.897	53963822	27652609	17.274	17.075
Target Compounds							
2)	A alpha-BHC	0.000	3.245	0	154866	N.D.	<MDL
3)	MA gamma-BHC...	0.000	3.634f	0	112636	N.D.	<MDL
6)	B beta-BHC	0.000	3.869	0	283084	N.D.	0.286
14)	MA Endrin	0.000	5.638	0	470355	N.D.	0.257
15)	B Endosulfa...	0.000	5.874f	0	410816	N.D.	0.227
23)	Toxaphene-2	0.000	5.638	0	470355	N.D.	39.005


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092123\
Data File : PD078144.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2023 13:45
Operator : AR\AJ
Sample : 04450-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 00:22:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 06 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.541 min
Delta R.T.: 0.000 min
Response: 53881586
Conc: 21.23 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.759 min
Delta R.T.: -0.001 min
Response: 33461235
Conc: 22.64 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.528 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 3.869 min
Delta R.T.: -0.020 min
Response: 283084
Conc: 0.29 ng/ml




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.587 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 5.638 min
Delta R.T.: 0.015 min
Response: 470355
Conc: 0.26 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.806 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 5.874 min
Delta R.T.: -0.043 min
Response: 410816
Conc: 0.23 ng/ml




#23 Toxaphene-2

R.T.: 0.000 min
Exp R.T.: 6.452 min
Response: 0
Conc: N.D.




#23 Toxaphene-2

R.T.: 5.638 min
Delta R.T.: -0.033 min
Response: 470355
Conc: 39.00 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.082 min
Delta R.T.: -0.003 min
Response: 53963822
Conc: 17.27 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.897 min
Delta R.T.: -0.004 min
Response: 27652609
Conc: 17.07 ng/ml