

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD092218CLP.M
 Title : GC Extractables
 Last Update : Sat Sep 22 00:14:15 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049335.D 1 =PD049333.D 2 =PD049334.D
 4 =PD049336.D 5 =PD049337.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.027	1.068	1.042	1.036	1.046	1.044	E6	1.47
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	8.073	8.068	7.922	8.248	8.731	8.208	E3	3.82
23)	Toxaphene-2	8.985	8.950	8.799	9.017	9.318	9.014	E3	2.10
24)	Toxaphene-3	2.353	2.228	2.276	2.383	2.486	2.345	E4	4.24
25)	Toxaphene-4	2.474	2.345	2.403	2.584	2.702	2.501	E4	5.72
26)	Toxaphene-5	3.015	2.874	2.884	3.022	3.176	2.994	E4	4.13
27) SA	Decachlorobiphenyl	1.212	1.224	1.219	1.192	1.209	1.211	E6	1.00

Signal #2 Calibration Files

3 =PD049335.D 1 =PD049333.D 2 =PD049334.D
 4 =PD049336.D 5 =PD049337.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	9.477	9.714	9.736	9.219	8.921	9.413	E6	3.68
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD092218CLP.M
 Title : GC Extractables
 Last Update : Sat Sep 22 00:14:15 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049335.D 1 =PD049333.D 2 =PD049334.D
 4 =PD049336.D 5 =PD049337.D

Compound		3	1	2	4	5	Avg	%RSD	
21) B	Endrin ketone						0.000	-1.00	
22)	Toxaphene-1	4.401	4.546	4.473	4.261	4.139	4.364	E4	3.76
23)	Toxaphene-2	6.346	6.584	6.412	6.295	6.160	6.359	E4	2.45
24)	Toxaphene-3	2.501	2.472	2.509	2.444	2.431	2.471	E5	1.38
25)	Toxaphene-4	4.380	4.140	4.350	4.390	4.699	4.392	E4	4.55
26)	Toxaphene-5	7.268	7.514	6.933	6.911	6.997	7.125	E4	3.65
27) SA	Decachlorobiphenyl	4.294	4.460	4.366	4.267	4.179	4.314	E6	2.45

(#) = Out of Range