

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD092218CLP.M
 Title : GC Extractables
 Last Update : Sat Sep 22 01:51:06 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049343.D 1 =PD049339.D 2 =PD049341.D
 4 =PD049345.D 5 =PD049347.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	0.962	1.010	0.970	0.948	0.994	0.977	E6	2.55
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin	1.374	1.359	1.341	1.393	1.492	1.392	E6	4.25
6) B	beta-BHC	6.437	6.897	6.642	6.234	6.399	6.522	E5	3.91
7) B	delta-BHC	1.502	1.447	1.449	1.537	1.661	1.520	E6	5.78
8) B	Heptachlor epoxide	1.323	1.340	1.310	1.323	1.399	1.339	E6	2.62
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane	1.353	1.377	1.340	1.355	1.447	1.374	E6	3.10
11) B	cis-Chlordane	1.311	1.346	1.303	1.303	1.382	1.329	E6	2.59
12) B	4,4'-DDE	1.329	1.259	1.284	1.363	1.485	1.344	E6	6.57
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II	1.280	1.288	1.273	1.268	1.335	1.289	E6	2.07
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde	1.040	1.071	1.036	1.027	1.067	1.048	E6	1.89
19) B	Endosulfan Sulfate	1.198	1.189	1.185	1.182	1.238	1.198	E6	1.90
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone	1.359	1.330	1.345	1.357	1.391	1.357	E6	1.65
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00
24)	Toxaphene-3						0.000		-1.00
25)	Toxaphene-4						0.000		-1.00
26)	Toxaphene-5						0.000		-1.00
27) SA	Decachlorobiphenyl	1.090	1.114	1.109	1.071	1.059	1.088	E6	2.19

Signal #2 Calibration Files

3 =PD049343.D 1 =PD049339.D 2 =PD049341.D
 4 =PD049345.D 5 =PD049347.D

	Compound	3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	9.261	9.463	9.273	8.763	8.657	9.083	E6	3.88
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin	1.122	1.162	1.130	1.069	1.043	1.105	E7	4.35
6) B	beta-BHC	5.143	5.434	5.195	4.838	4.756	5.073	E6	5.45
7) B	delta-BHC	1.132	1.113	1.121	1.086	1.088	1.108	E7	1.85
8) B	Heptachlor epoxide	1.012	1.060	1.020	0.958	0.930	0.996	E7	5.21
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane	0.984	1.014	0.996	0.939	0.924	0.971	E7	3.94
11) B	cis-Chlordane	9.389	9.928	9.625	8.933	8.745	9.324	E6	5.22
12) B	4,4'-DDE	7.262	7.548	7.253	6.839	6.899	7.160	E6	4.07
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II	7.554	8.227	7.712	7.084	6.981	7.512	E6	6.71
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde	5.597	5.745	5.640	5.296	5.205	5.497	E6	4.24
19) B	Endosulfan Sulfate	5.911	6.036	5.972	5.583	5.586	5.818	E6	3.74
20) A	Methoxychlor						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD092218CLP.M
Title : GC Extractables
Last Update : Sat Sep 22 01:51:06 2018
Response Via : Initial Calibration

Calibration Files

3	=PD049343.D	1	=PD049339.D	2	=PD049341.D
4	=PD049345.D	5	=PD049347.D		

	Compound	3	1	2	4	5	Avg		%RSD
21) B	Endrin ketone	6.079	6.205	6.208	5.825	5.716	6.006	E6	3.75
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00
24)	Toxaphene-3						0.000		-1.00
25)	Toxaphene-4						0.000		-1.00
26)	Toxaphene-5						0.000		-1.00
27) SA	Decachlorobiphenyl	4.174	4.419	4.291	3.956	3.759	4.120	E6	6.40

(#) = Out of Range