

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092218\
 Data File : PD049344.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Sep 2018 22:46
 Operator : AJ\SJ
 Sample : INDA401
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA401

Manual Integrations
 APPROVED

Sohil
 9/24/2018 11:56:42 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 22 01:01:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092218CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 22 01:00:41 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.296	3.913	39744687	366.8E6	40.134	40.151
27) SA Decachlor...	7.954	8.904	87968166	326.4E6	80.829	79.233
Target Compounds						
2) A alpha-BHC	3.722	4.291	64105179	533.9E6	40.170	40.179
3) MA gamma-BHC...	3.996	4.569	60072290	490.0E6	40.219	40.219
4) MA Heptachlor	4.279	5.075	59186989	391.8E6	40.373	40.254m
9) A Endosulfan I	5.281	6.108	52572850	359.5E6	40.457	40.529
13) MA Dieldrin	5.519	6.361	122.5E6	763.9E6	81.281	80.924
14) MA Endrin	5.772	6.575	100.4E6	570.3E6	80.982	81.507
16) A 4,4'-DDD	5.906	6.698	94969929	366.2E6	81.211	80.565
17) MA 4,4'-DDT	6.141	6.996	101.4E6	416.4E6	81.250	80.485
20) A Methoxychlor	6.689	7.450	241.1E6	614.7E6	405.550	400.909

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092218\
Data File : PD049344.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Sep 2018 22:46
Operator : AJ\SJ
Sample : INDA401
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :

ECD_D

Client Sample Id :

INDA401

Manual Integrations

APPROVED

Sohil

9/24/2018 11:56:42 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 22 01:01:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092218CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 22 01:00:41 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

