

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD092224\  
 Data File : PD085520.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 23 Sep 2024 05:48  
 Operator : AR\AJ  
 Sample : INDA334  
 Misc :  
 ALS Vial : 47 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 INDA3230

Manual Integrations  
 APPROVED

Reviewed By : Abdul Mirza 09/25/2024  
 Supervised By : Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Sep 24 09:21:32 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD091924CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Sep 20 05:48:22 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm


|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml   | ng/ml    |
|-----------------------------|-----------------|-------|-------|----------|----------|---------|----------|
| -----                       |                 |       |       |          |          |         |          |
| System Monitoring Compounds |                 |       |       |          |          |         |          |
| 1)                          | SA Tetrachlo... | 3.487 | 2.892 | 24272003 | 83213438 | 20.043  | 22.015   |
| 27)                         | SA Decachlor... | 8.959 | 8.084 | 70535173 | 127.4E6  | 42.997  | 47.668   |
| Target Compounds            |                 |       |       |          |          |         |          |
| 2)                          | A alpha-BHC     | 3.931 | 3.402 | 49603560 | 120.6E6  | 20.994  | 21.848   |
| 3)                          | MA gamma-BHC... | 4.258 | 3.737 | 78584398 | 113.0E6  | 31.411  | 21.687 # |
| 4)                          | MA Heptachlor   | 4.853 | 4.089 | 44531033 | 103.9E6  | 21.089  | 21.886   |
| 9)                          | A Endosulfan I  | 5.992 | 5.253 | 39846835 | 86513802 | 22.482m | 21.053m  |
| 13)                         | MA Dieldrin     | 6.266 | 5.521 | 90048429 | 209.9E6  | 46.578  | 45.088   |
| 14)                         | MA Endrin       | 6.491 | 5.799 | 71615944 | 178.3E6  | 44.543m | 42.977   |
| 16)                         | A 4,4'-DDD      | 6.624 | 5.938 | 67346182 | 170.3E6  | 51.115m | 50.851   |
| 17)                         | MA 4,4'-DDT     | 6.939 | 6.193 | 49649281 | 130.5E6  | 40.729m | 41.186   |
| 20)                         | A Methoxychlor  | 7.413 | 6.765 | 142.3E6  | 332.6E6  | 194.558 | 204.914  |
| -----                       |                 |       |       |          |          |         |          |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD092224\  
Data File : PD085520.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 23 Sep 2024 05:48  
Operator : AR\AJ  
Sample : INDA334  
Misc :  
ALS Vial : 47 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Sep 24 09:21:32 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD091924CLP.M  
Quant Title : GC Extractables  
QLast Update : Fri Sep 20 05:48:22 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Miza 09/25/2024  
Supervised By: Ankita Jodhani 09/25/2024

Manual Integrations  
APPROVED

Instrument :  
ECD\_D  
Client Sampled :  
INDA3230