

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092320\
 Data File : PD059308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2020 17:06
 Operator : DD\AJ
 Sample : L1955-09
 Misc : PEST MDL 1 PPB
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 01:21:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092320CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 23 13:19:45 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.206	3.865	19671541	36934299	20.401	20.655
27)	SA Decachlor...	7.840	8.862	39402618	59348762	38.607	40.256
Target Compounds							
2)	A alpha-BHC	3.628	4.248	1432230	2618177	0.998	1.027
3)	MA gamma-BHC...	3.899	4.528	2167667	2121628	1.548	0.907 #
4)	MA Heptachlor	4.178	5.028	1280677	2424069	1.039	1.075
5)	MB Aldrin	4.412	5.328	1242134	2071606	0.995	0.973
6)	B beta-BHC	4.148	4.701	668723	956529	1.123	0.915
7)	B delta-BHC	4.338	4.909	1322278	526806	0.967	0.289m#
8)	B Heptachlo...	4.845	5.711	1202463	3045322	1.045	1.532 #
9)	A Endosulfan I	5.171	6.065	1101072	1910219	1.009	1.046
10)	B trans-Chl...	5.064	5.945	1183287	2125749	1.018	1.070
11)	B cis-Chlor...	5.121	6.017	1224895	2105647	1.050	1.056
12)	B 4,4'-DDE	5.290	6.177	1140329	1895920	0.993	1.004
13)	MA Dieldrin	5.408	6.320	1185968	1958666	0.994	0.981
14)	MA Endrin	5.660	6.535	917005	1439272	1.033	0.966
15)	B Endosulfa...	5.933	6.746	1251562	2071101	1.230	1.236
16)	A 4,4'-DDD	5.797	6.663	913340	1506861	1.015	0.964m
17)	MA 4,4'-DDT	6.029	6.960	643911	1420633	0.709m	0.915m#
18)	B Endrin al...	6.100	6.870	746315	1717054	0.888	1.188 #
19)	B Endosulfa...	6.311	7.092	1187883	1839015	1.169	1.069
20)	A Methoxychlor	6.577	7.421	395234	795188	0.916m	0.966
21)	B Endrin ke...	6.791	7.558	1343638	2014680	1.210	1.165

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092320\
Data File : PD059308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2020 17:06
Operator : DD\AJ
Sample : L1955-09
Misc : PEST MDL 1 PPB
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:21:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092320CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 23 13:19:45 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSampled :