

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.448min 1478.683 ng/ml

response 1790654543

(1) Tetrachloro-m-xylene #2 (SA)

2.906min -1052.497 ng/ml

response -3978286405

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.488min 67.800 ng/ml m

response 82104053

(1) Tetrachloro-m-xylene #2 (SA)

2.905min 502.846 ng/ml m

response 1900686622

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)

6.629min 7.662 ng/ml m
response 10094871

(16) 4,4'-DDD #2 (A)

5.941min 31.760 ng/ml m
response 106359025

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) cis-Chlordane (B)
5.941min 112.855 ng/ml
response 226237234

(11) cis-Chlordane #2 (B)
5.203min 89.392 ng/ml
response 414697024

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(11) cis-Chlordane (B)
5.939min 112.530 ng/ml m
response 225586809

(11) cis-Chlordane #2 (B)
5.202min 103.529 ng/ml m
response 480280836

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.133min 48.845 ng/ml

response 85386289

(12) 4,4'-DDE #2 (B)

5.427min 29.628 ng/ml

response 126937880

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.131min 41.266 ng/ml m

response 72136752

(12) 4,4'-DDE #2 (B)

5.376min 38.247 ng/ml m

response 163864938

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) 4,4'-DDD (A)
6.635min 164.902 ng/ml
response 217267313

(16) 4,4'-DDD #2 (A)
5.942min 12.109 ng/ml
response 40552167

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

6.942min 407.873 ng/ml

response 497203205

(17) 4,4'-DDT #2 (MA)

6.192min 274.114 ng/ml

response 868232976

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(17) 4,4'-DDT (MA)

6.940min 506.327 ng/ml m
response 617220969

(17) 4,4'-DDT #2 (MA)

6.192min 274.114 ng/ml
response 868232976

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(20) Methoxychlor (A)

7.406min 129.696 ng/ml

response 94891535

(20) Methoxychlor #2 (A)

6.752min 47.787 ng/ml

response 77555371

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(20) Methoxychlor (A)
7.405min 202.218 ng/ml m
response 147952143

(20) Methoxychlor #2 (A)
6.751min 359.311 ng/ml m
response 583135767

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DDC95RE

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024
Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

8.950min 15.429 ng/ml

response 25309791

(27) Decachlorobiphenyl #2 (SA)

8.131min 93.294 ng/ml

response 249402603

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092324\
Data File : PD085562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2024 22:15
Operator : AR\AJ
Sample : P4017-13RE
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

DDC95RE


Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/25/2024

Supervised By :Ankita Jodhani 09/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 01:43:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(27) Decachlorobiphenyl (SA)

8.949min 45.480 ng/ml m

response 74607675

(27) Decachlorobiphenyl #2 (SA)

8.080min 3.704 ng/ml m

response 9902588