

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092421\
 Data File : PD065846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2021 11:56
 Operator : AR\AJ
 Sample : INDA307
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3153

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 04:14:29 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092021CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Oct 04 09:46:55 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.495	4.012	16516514	134.3E6	19.083	18.892
27)	SA Decachlor...	8.240	9.086	54764735	164.4E6	37.036	38.019
Target Compounds							
2)	A alpha-BHC	3.935	4.407	22092423	178.6E6	18.676	18.783
3)	MA gamma-BHC...	4.219	4.693	22653390	169.1E6	18.532	18.500
4)	MA Heptachlor	4.516	5.197	24101654	155.9E6	18.865	18.526
9)	A Endosulfan I	5.546	6.250	22958629	121.1E6	19.237	18.568
13)	MA Dieldrin	5.789	6.509	48828745	254.0E6	37.732	37.192
14)	MA Endrin	6.047	6.728	38826234	186.3E6	33.020	33.394
16)	A 4,4'-DDD	6.169	6.852	40603760	202.7E6	37.971	36.703
17)	MA 4,4'-DDT	6.409	7.149	44531857	201.2E6	38.399	37.192
20)	A Methoxychlor	6.956	7.609	129.7E6	480.1E6	188.639	183.833

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092421\
Data File : PD065846.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2021 11:56
Operator : AR\AJ
Sample : INDA307
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 04:14:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092021CLP.M
Quant Title : GC Extractables
QLast Update : Mon Oct 04 09:46:55 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

