

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092424\
 Data File : PD085600.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2024 10:46
 Operator : AR\AJ
 Sample : P4017-02DL 5X
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDC41DL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/26/2024
 Supervised By : Ankita Jodhani 09/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 05:10:08 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 20 05:48:22 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.483	2.901	4070013	19322666	3.361m	5.112m#
2) SA Decachlor...	8.946	8.078	7601375	4857297	4.634m	1.817m#
Target Compounds						
2) A alpha-BHC	3.931	3.401	110.3E6	231.1E6	46.692	41.845
7) B delta-BHC	4.689	4.268	154.3E6	254.4E6	67.704	48.301 #
9) A Endosulfan I	5.995	5.271	10339514	8223479	5.834m	2.001m#
11) B cis-Chlor...	5.936	5.202	9677838	30597216	4.828	6.596 #
13) MA Dieldrin	6.266	5.518	23881526	12861762	12.353m	2.763m#
16) A 4,4'-DDD	6.623	5.938	35267629	70872080	26.768m	21.163
17) MA 4,4'-DDT	6.938	6.192	111.0E6	276.3E6	91.050m	87.221
21) B Endrin ke...	7.521	6.994	51256420	16037869	32.061m	3.453m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092424\
Data File : PD085600.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2024 10:46
Operator : AR\AJ
Sample : P4017-02DL 5X
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 05:10:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

