

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092424\
 Data File : PD085611.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2024 13:14
 Operator : AR\AJ
 Sample : P4017-17DL 3X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDCD3DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/07/2024
 Supervised By :Ankita Jodhani 10/07/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 05:14:33 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 20 05:48:22 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.475	2.902	11700687	997.4E6	9.662m	263.866m#
27)	SA Decachlor...	8.948	8.081	26164638	13239020	15.950m	4.952m#
Target Compounds							
2)	A alpha-BHC	3.931	3.401	37243478	71933572	15.763	13.027
7)	B delta-BHC	4.689	4.266	150.4E6	200.6E6	65.955m	38.080m#
11)	B cis-Chlor...	5.934	5.200	18011897	45055746	8.985m	9.712m
12)	B 4,4'-DDE	6.119	5.386	26584831	33639170	15.208m	7.852m#
13)	MA Dieldrin	6.266	5.516	19983153	21537651	10.336m	4.627m#
16)	A 4,4'-DDD	6.624	5.938	157.3E6	262.1E6	119.425m	78.255 #
17)	MA 4,4'-DDT	6.946	6.188	43591698	33568516	35.760m	10.598m#
20)	A Methoxychlor	7.422	6.749	20455100	226.5E6	27.958m	139.532m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092424\
Data File : PD085611.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2024 13:14
Operator : AR\AJ
Sample : P4017-17DL 3X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 05:14:33 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091924CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 20 05:48:22 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By : Yogesh Patel 10/07/2024
Supervised By : Ankita Jodhani 10/07/2024

Instrument :
ECD_D
ClientSampled :
DDCD3DL