

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092519\
 Data File : PD054954.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Sep 2019 10:29
 Operator : SG\AJ
 Sample : MDL-S-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 MDL-S-05

Manual Integrations
 APPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 10:48:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092319CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 24 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.295	3.966	14820561	20083249	19.481	19.570
27)	SA Decachlor...	7.975	9.003	28987430	38676368	34.403	33.197
Target Compounds							
2)	A alpha-BHC	3.722	4.353	1403359	1198901	1.264	0.832 #
3)	MA gamma-BHC...	3.998	4.635	1351530	1405648	1.260	1.004
4)	MA Heptachlor	4.283	5.142	885767	1524838	0.958	1.082
5)	MB Aldrin	4.523	5.446	887392	1209951	0.912	0.915
6)	B beta-BHC	4.246	4.802	421001	679183	0.900	1.019m
7)	B delta-BHC	4.442	5.019	917109	1311565	0.857	0.904
8)	B Heptachlo...	4.961	5.830	899902	1450649	0.981	1.103
9)	A Endosulfan I	5.292	6.135f	897813	1452751	1.012	1.151
10)	B trans-Chl...	5.182	6.063	969713	1508493	1.027	1.113
11)	B cis-Chlor...	5.240	6.135	966559	1452751	1.056	1.103
12)	B 4,4'-DDE	5.409	6.291	804844	1236963	0.897	0.956
13)	MA Dieldrin	5.532	6.441	984392	1419332	1.000	1.017
14)	MA Endrin	5.785	6.655	917255	1140060	1.103	1.035m
15)	B Endosulfa...	6.056	6.864	842263	1255907	0.967m	0.988m
16)	A 4,4'-DDD	5.920	6.778	702580	1090361	0.969	1.005m
17)	MA 4,4'-DDT	6.157	7.076	216362	891254	0.325	0.814m#
18)	B Endrin al...	6.227	6.987	766982	1107588	1.055m	1.069m
19)	B Endosulfa...	6.439	7.209	892279	1259502	1.099	1.060m
20)	A Methoxychlor	6.706	7.532	245950	564153	0.694m	0.904m#
21)	B Endrin ke...	6.921	7.679	625791	1371436	0.693	1.041 #

Ankita

9/26/2019 3:45:25 PM

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092519\
Data File : PD054954.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Sep 2019 10:29
Operator : SG\AJ
Sample : MDL-S-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
MDL-S-05

Manual Integrations
APPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 10:48:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092319CLP.M
Quant Title : GC Extractables
QLast Update : Tue Sep 24 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

