

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092523\
 Data File : PD078266.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Sep 2023 10:27
 Operator : AR\AJ
 Sample : PB155764BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PB155764BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 00:20:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090323.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 01 16:18:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.546	2.757	34902776	22032050	13.746	13.930
2)	SA Decachlor...	9.088	7.897	46455284	27387393	15.250	15.871
Target Compounds							
2)	A alpha-BHC	4.003	3.259	185.2E6	105.4E6	45.025	45.579
3)	MA gamma-BHC...	4.338	3.588	169.2E6	98506183	42.770	44.065
4)	MA Heptachlor	4.928	3.927	183.9E6	106.4E6	46.936	47.309
5)	MB Aldrin	5.270	4.207	177.8E6	100.7E6	44.199	44.188
6)	B beta-BHC	4.535	3.886	68194420	40441839	42.257	45.376
7)	B delta-BHC	4.783	4.116	164.0E6	97405584	44.487	44.932
8)	B Heptachlo...	5.698	4.709	155.7E6	93328786	43.532	46.015
9)	A Endosulfan I	6.085	5.079	148.8E6	89530280	45.249	47.797
10)	B gamma-Chl...	5.955	4.960	159.2E6	94549686	45.210	47.035
11)	B alpha-Chl...	6.034	5.024	157.7E6	93900091	44.331	46.607
12)	B 4,4'-DDE	6.207	5.212	152.5E6	90189888	47.356	49.411
13)	MA Dieldrin	6.360	5.344	154.1E6	94263247	43.309	46.167
14)	MA Endrin	6.591	5.620	131.9E6	84466112	43.494	46.632
15)	B Endosulfa...	6.811	5.915	125.9E6	76818807	40.643	44.896
16)	A 4,4'-DDD	6.725	5.768	116.8E6	70311306	45.575	50.315
17)	MA 4,4'-DDT	7.040	6.019	121.4E6	71289437	46.416	50.458
18)	B Endrin al...	6.941	6.095	99133902	61401190	41.716	45.488
19)	B Endosulfa...	7.176	6.318	121.6E6	76934895	42.866	48.047
20)	A Methoxychlor	7.516	6.595	62792005	35923694	40.596	46.399
21)	B Endrin ke...	7.662	6.824	133.9E6	87023832	41.202	46.026
22)	Mirex	8.137	7.005	97639430	69086480	40.598	45.117

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092523\
Data File : PD078266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Sep 2023 10:27
Operator : AR\AJ
Sample : PB155764BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PB155764BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 00:20:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090323.M
Quant Title : GC Extractables
QLast Update : Fri Sep 01 16:18:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

