

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092721\
 Data File : PD065906.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2021 15:11
 Operator : AR\AJ
 Sample : PCHLORCCC500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PCHLORCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 16:34:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092221.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 23 12:18:33 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.495	4.012	44362905	291.7E6	59.905	50.803
28)	SA Decachlor...	8.240	9.088	57219128	184.6E6	41.890	47.704
Target Compounds							
23)	Chlordane-1	4.365	5.008	15309805	104.8E6	457.048	473.266
24)	Chlordane-2	4.866	5.478	19853433	106.6E6	488.861	460.774
25)	Chlordane-3	5.433	6.126	55411527	363.9E6	470.329	443.741
26)	Chlordane-4	5.491	6.199	46437997	450.4E6	470.190	469.605
27)	Chlordane-5	6.327	7.010	21959109	88817353	451.832	477.924



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092721\
Data File : PD065906.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2021 15:11
Operator : AR\AJ
Sample : PCHLORCCC500
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 16:34:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092221.M
Quant Title : GC Extractables
QLast Update : Thu Sep 23 12:18:33 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.495 min
Delta R.T.: 0.002 min
Response: 44362905
Conc: 59.90 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#1 Tetrachloro-m-xylene

R.T.: 4.012 min
Delta R.T.: 0.000 min
Response: 291720798
Conc: 50.80 ng/ml




#23 Chlordane-1

R.T.: 4.365 min
Delta R.T.: 0.003 min
Response: 15309805
Conc: 457.05 ng/ml



#23 Chlordane-1


R.T.: 5.008 min
Delta R.T.: 0.000 min
Response: 104787459
Conc: 473.27 ng/ml



#24 Chlordane-2


R.T.: 4.866 min
Delta R.T.: 0.003 min
Response: 19853433
Conc: 488.86 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#24 Chlordane-2

R.T.: 5.478 min
Delta R.T.: 0.000 min
Response: 106582616
Conc: 460.77 ng/ml




#25 Chlordane-3

R.T.: 5.433 min
Delta R.T.: 0.004 min
Response: 55411527
Conc: 470.33 ng/ml



#25 Chlordane-3


R.T.: 6.126 min
Delta R.T.: 0.000 min
Response: 363892427
Conc: 443.74 ng/ml



#26 Chlordane-4


R.T.: 5.491 min
Delta R.T.: 0.004 min
Response: 46437997
Conc: 470.19 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500




#26 Chlordane-4

R.T.: 6.199 min
Delta R.T.: 0.002 min
Response: 450427255
Conc: 469.61 ng/ml




#27 Chlordane-5

R.T.: 6.327 min
Delta R.T.: 0.005 min
Response: 21959109
Conc: 451.83 ng/ml



#27 Chlordane-5


R.T.: 7.010 min
Delta R.T.: 0.002 min
Response: 88817353
Conc: 477.92 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.240 min
Delta R.T.: 0.006 min
Response: 57219128
Conc: 41.89 ng/ml

Instrument :
ECD_D
ClientSampleId :
PCHLORCCC500



#28 Decachlorobiphenyl

R.T.: 9.088 min
Delta R.T.: 0.003 min
Response: 184623643
Conc: 47.70 ng/ml