

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(22) Toxaphene-1
5.418min -40.763 ng/ml
response -640573

(22) Toxaphene-1 #2
5.887min 145.294 ng/ml
response 8386505

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(22) Toxaphene-1

5.388min 124.243 ng/ml m

response 1952428

(22) Toxaphene-1 #2

5.887min 145.294 ng/ml

response 8386505

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(23) Toxaphene-2
5.750min 138.693 ng/ml
response 2328728

(23) Toxaphene-2 #2
6.268min 116.558 ng/ml
response 8591965

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(23) Toxaphene-2
5.749min 112.944 ng/ml m
response 1896394

(23) Toxaphene-2 #2
6.268min 116.558 ng/ml
response 8591965

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(24) Toxaphene-3
6.035min 43.599 ng/ml
response 831315

(24) Toxaphene-3 #2
7.093min 130.440 ng/ml
response 24255338

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(24) Toxaphene-3
6.033min 121.092 ng/ml m
response 2308877

(24) Toxaphene-3 #2
7.093min 130.440 ng/ml
response 24255338

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(25) Toxaphene-4
6.799min 136.063 ng/ml
response 10660303

(25) Toxaphene-4 #2
7.186min 67.881 ng/ml
response 17246489

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(25) Toxaphene-4
6.797min 99.501 ng/ml m
response 7795716

(25) Toxaphene-4 #2
7.184min 129.994 ng/ml m
response 33027752

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
 Data File : PD072058.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Sep 2022 11:36
 Operator : AR\AJ
 Sample : N4519-02
 Misc : TOX MDL 100 PPB
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 ECD_D
 ClientSampleId :
 MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
 Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 27 23:46:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 23 17:46:27 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(26) Toxaphene-5
 7.116min 98.133 ng/ml
 response 3282030

(26) Toxaphene-5 #2
 7.615min 210.938 ng/ml
 response 31753486

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092722\
Data File : PD072058.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2022 11:36
Operator : AR\AJ
Sample : N4519-02
Misc : TOX MDL 100 PPB
ALS Vial : 9 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
MDL-WATER-QT4-2022-08

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/28/2022
Supervised By :Ankita Jodhani 09/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 23:46:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092422CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 23 17:46:27 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(26) Toxaphene-5

7.116min 98.133 ng/ml

response 3282030

(26) Toxaphene-5 #2

7.613min 134.493 ng/ml m

response 20245830