

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092723\
 Data File : PD078313.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Sep 2023 11:34
 Operator : AR\AJ
 Sample : 04480-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 26 21:36:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 06 16:40:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.758	52890521	34478326	20.842	23.330
27)	SA Decachlor...	9.079	7.895	66566120	38310994	21.309	23.656
Target Compounds							
4)	MA Heptachlor	0.000	3.941	0	1057656	N.D.	0.476
5)	MB Aldrin	0.000	4.199	0	576048	N.D.	0.257
6)	B beta-BHC	0.000	3.841f	0	383261	N.D.	0.388
8)	B Heptachlo...	0.000	4.665f	0	128223	N.D.	<MDL
9)	A Endosulfan I	0.000	5.032f	0	598553	N.D.	0.333
11)	B cis-Chlor...	0.000	5.032	0	598553	N.D.	0.293
12)	B 4,4'-DDE	0.000	5.247	0	440363	N.D.	0.232
15)	B Endosulfa...	0.000	5.872f	0	372054	N.D.	0.206
21)	B Endrin ke...	0.000	6.826	0	347047	N.D.	0.181
22)	Toxaphene-1	0.000	5.032f	0	598553	N.D.	57.345


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092723\
Data File : PD078313.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Sep 2023 11:34
Operator : AR\AJ
Sample : 04480-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 26 21:36:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 06 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.540 min
Delta R.T.: -0.002 min
Response: 52890521
Conc: 20.84 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.758 min
Delta R.T.: -0.003 min
Response: 34478326
Conc: 23.33 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 4.923 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 3.941 min
Delta R.T.: 0.010 min
Response: 1057656
Conc: 0.48 ng/ml



#5 Aldrin


R.T.: 0.000 min
Exp R.T.: 5.264 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#5 Aldrin

R.T.: 4.199 min
Delta R.T.: -0.011 min
Response: 576048
Conc: 0.26 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T.: 4.528 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 3.841 min
Delta R.T.: -0.047 min
Response: 383261
Conc: 0.39 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 6.081 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#9 Endosulfan I

R.T.: 5.032 min
Delta R.T.: -0.051 min
Response: 598553
Conc: 0.33 ng/ml




#11 cis-Chlordane

R.T.: 0.000 min
Exp R.T.: 6.029 min
Response: 0
Conc: N.D.



#11 cis-Chlordane


R.T.: 5.032 min
Delta R.T.: 0.006 min
Response: 598553
Conc: 0.29 ng/ml



#12 4,4'-DDE


R.T.: 0.000 min
Exp R.T.: 6.201 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#12 4,4'-DDE

R.T.: 5.247 min
Delta R.T.: 0.032 min
Response: 440363
Conc: 0.23 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 6.806 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 5.872 min
Delta R.T.: -0.045 min
Response: 372054
Conc: 0.21 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T.: 7.658 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#21 Endrin ketone

R.T.: 6.826 min
Delta R.T.: 0.000 min
Response: 347047
Conc: 0.18 ng/ml




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T.: 6.247 min
Response: 0
Conc: N.D.



#22 Toxaphene-1


R.T.: 5.032 min
Delta R.T.: 0.044 min
Response: 598553
Conc: 57.34 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.079 min
Delta R.T.: -0.005 min
Response: 66566120
Conc: 21.31 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 7.895 min
Delta R.T.: -0.006 min
Response: 38310994
Conc: 23.66 ng/ml