

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092723\
Data File : PD078394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 06:16
Operator : AR\AJ
Sample : 04588-09
Misc :
ALS Vial : 80 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BBHT7

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/27/2023
Supervised By :Ankita Jodhani 09/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 07:38:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 06 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.539min 12.986 ng/ml m

response 32953875

(1) Tetrachloro-m-xylene #2 (SA)

2.759min 11.322 ng/ml m

response 16733289

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092723\
Data File : PD078394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 06:16
Operator : AR\AJ
Sample : 04588-09
Misc :
ALS Vial : 80 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BBHT7

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/27/2023
Supervised By :Ankita Jodhani 09/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 01:58:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 06 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.201min 30.894 ng/ml

response 109525664

(12) 4,4'-DDE #2 (B)

5.207min 15.329 ng/ml

response 29119004

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092723\
Data File : PD078394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Sep 2023 06:16
Operator : AR\AJ
Sample : 04588-09
Misc :
ALS Vial : 80 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
BBHT7

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 09/27/2023
Supervised By :Ankita Jodhani 09/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 27 07:38:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 06 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(12) 4,4'-DDE (B)

6.200min 27.979 ng/ml m

response 99188376

(12) 4,4'-DDE #2 (B)

5.205min 17.143 ng/ml m

response 32565022