

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD092818CLP.M
 Title : GC Extractables
 Last Update : Sat Sep 29 01:00:01 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049484.D 1 =PD049482.D 2 =PD049483.D
 4 =PD049485.D 5 =PD049486.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	9.038	8.952	9.006	8.592	8.705	8.858 E5	2.24
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	5.652	5.569	5.477	5.301	5.706	5.541 E3	2.88
23)	Toxaphene-2	6.956	7.253	7.126	7.001	7.022	7.071 E3	1.69
24)	Toxaphene-3	1.802	1.811	1.787	1.831	1.945	1.835 E4	3.45
25)	Toxaphene-4	1.851	1.757	1.776	1.859	1.959	1.840 E4	4.34
26)	Toxaphene-5	1.999	1.934	1.961	2.015	2.138	2.010 E4	3.91
27) SA	Decachlorobiphenyl	8.155	8.804	8.438	7.668	7.404	8.094 E5	7.00

Signal #2 Calibration Files

3 =PD049484.D 1 =PD049482.D 2 =PD049483.D
 4 =PD049485.D 5 =PD049486.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.207	1.202	1.221	1.136	1.088	1.171 E7	4.84
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1	5.189	5.371	5.566	4.936	4.584	5.129 E4	7.47
23)	Toxaphene-2	5.913	6.606	5.835	5.523	5.515	5.878 E4	7.56

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD092818CLP.M
 Title : GC Extractables
 Last Update : Sat Sep 29 01:00:01 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049484.D 1 =PD049482.D 2 =PD049483.D
 4 =PD049485.D 5 =PD049486.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	2.488	2.604	2.442	2.394	2.448	2.475	E5	3.21
25)	Toxaphene-4	3.772	3.773	3.692	4.001	4.047	3.857	E4	4.06
26)	Toxaphene-5	7.064	6.699	7.004	7.429	7.471	7.134	E4	4.50
27) SA	Decachlorobiphenyl	5.398	5.499	5.512	5.223	5.069	5.340	E6	3.57

(#) = Out of Range