

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD092818CLP.M
 Title : GC Extractables
 Last Update : Sat Sep 29 01:34:45 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049491.D 1 =PD049487.D 2 =PD049489.D
 4 =PD049493.D 5 =PD049495.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	8.645	8.881	8.718	8.672	8.598	8.703	E5	1.25
2) A	alpha-BHC	1.205	1.137	1.164	1.266	1.310	1.216	E6	5.87
3) MA	gamma-BHC (Lindane)	1.190	1.181	1.170	1.215	1.238	1.199	E6	2.27
4) MA	Heptachlor	1.253	1.264	1.246	1.280	1.295	1.267	E6	1.57
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	1.094	1.139	1.109	1.092	1.087	1.104	E6	1.92
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	1.181	1.153	1.164	1.232	1.263	1.199	E6	3.89
14) MA	Endrin	9.613	9.319	9.331	9.783	9.982	9.606	E5	2.99
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD	7.654	7.582	7.580	7.900	8.065	7.756	E5	2.79
17) MA	4,4'-DDT	7.493	7.203	7.331	7.777	8.118	7.584	E5	4.84
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	4.086	3.959	4.102	4.198	4.304	4.130	E5	3.13
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00
24)	Toxaphene-3						0.000		-1.00
25)	Toxaphene-4						0.000		-1.00
26)	Toxaphene-5						0.000		-1.00
27) SA	Decachlorobiphenyl	7.808	8.498	8.158	7.457	7.147	7.813	E5	6.89

Signal #2 Calibration Files

3 =PD049491.D 1 =PD049487.D 2 =PD049489.D
 4 =PD049493.D 5 =PD049495.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.206	1.247	1.227	1.163	1.107	1.190	E7	4.70
2) A	alpha-BHC	1.678	1.650	1.681	1.663	1.612	1.657	E7	1.69
3) MA	gamma-BHC (Lindane)	1.545	1.607	1.564	1.512	1.459	1.538	E7	3.61
4) MA	Heptachlor	1.457	1.496	1.489	1.430	1.376	1.450	E7	3.38
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	1.107	1.136	1.112	1.074	1.030	1.092	E7	3.76
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	1.147	1.172	1.173	1.120	1.090	1.140	E7	3.12
14) MA	Endrin	7.480	7.657	7.328	7.265	7.082	7.362	E6	2.96
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD	6.857	7.211	6.871	6.797	6.638	6.875	E6	3.04
17) MA	4,4'-DDT	4.580	4.293	4.515	5.093	5.196	4.735	E6	8.24
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	2.430	2.206	2.363	2.485	2.599	2.417	E6	6.04
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD092818CLP.M
 Title : GC Extractables
 Last Update : Sat Sep 29 01:34:45 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049491.D 1 =PD049487.D 2 =PD049489.D
 4 =PD049493.D 5 =PD049495.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	5.352	5.328	5.431	5.252	5.031	5.279 E6	2.89

(#) = Out of Range