

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092818\
 Data File : PD049485.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Sep 2018 20:54
 Operator : AJ\SJ
 Sample : TOXAPH401
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 29 00:46:04 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092818CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 29 00:28:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.272	4.060	34366121	454.5E6	39.150	39.736
27)	SA Decachlor...	7.879	9.170	61345807	417.8E6	79.236	79.889
Target Compounds							
22)	Toxaphene-1	5.425	6.464	21205421	197.4E6	3818.768	4027.007
23)	Toxaphene-2	5.572	6.756	28002315	220.9E6	4004.536	3909.890
24)	Toxaphene-3	6.005	7.029	73220335	957.5E6	3938.461	3918.829
25)	Toxaphene-4	6.244	7.422	74347125	160.1E6	3934.740	4062.320
26)	Toxaphene-5	6.716	7.743	80614450	297.2E6	3930.388	4058.699

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD092818\
Data File : PD049485.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Sep 2018 20:54
Operator : AJ\SJ
Sample : TOXAPH401
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 29 00:46:04 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092818CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 29 00:28:23 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

