

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD093019.M
 Title : GC Extractables
 Last Update : Tue Oct 01 06:33:37 2019
 Response Via : Initial Calibration

Calibration Files

50 =PD055058.D 100 =PD055056.D 75 =PD055057.D
 25 =PD055059.D 5 =PD055060.D

	Compound	50	100	75	25	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	8.367	8.529	8.256	8.331	8.883	8.473 E5	2.95
2) MA	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	gamma-Chlordane						0.000	-1.00
11) B	alpha-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Mirex						0.000	-1.00
23)	Chlordane-1	2.439	2.531	2.459	2.606	2.293	2.466 E4	4.74
24)	Chlordane-2	3.050	2.942	2.926	3.160	3.123	3.040 E4	3.45
25)	Chlordane-3	9.135	9.167	8.954	9.088	9.473	9.164 E4	2.09
26)	Chlordane-4	8.004	8.068	7.772	7.968	8.306	8.023 E4	2.40
27)	Chlordane-5	2.955	2.881	2.829	2.893	2.786	2.869 E4	2.25
28) SA	Decachlorobiphenyl	0.850	0.810	0.805	0.867	1.006	0.868 E6	9.40

Signal #2 Calibration Files

50 =PD055058.D 100 =PD055056.D 75 =PD055057.D
 25 =PD055059.D 5 =PD055060.D

	Compound	50	100	75	25	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.110	1.113	1.087	1.121	1.214	1.129 E6	4.36
2) A	alpha-BHC						0.000	-1.00
3) MA	gamma-BHC (Lindane)						0.000	-1.00
4) MA	Heptachlor						0.000	-1.00
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I						0.000	-1.00
10) B	gamma-Chlordane						0.000	-1.00
11) B	alpha-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin						0.000	-1.00
14) MA	Endrin						0.000	-1.00
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD						0.000	-1.00
17) MA	4,4'-DDT						0.000	-1.00
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor						0.000	-1.00
21) B	Endrin ketone						0.000	-1.00
22)	Mirex						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD093019.M
 Title : GC Extractables
 Last Update : Tue Oct 01 06:33:37 2019
 Response Via : Initial Calibration

Calibration Files

50 =PD055058.D 100 =PD055056.D 75 =PD055057.D
 25 =PD055059.D 5 =PD055060.D

	Compound	50	100	75	25	5	Avg		%RSD
23)	Chlordane-1	4.312	4.279	4.211	4.430	5.097	4.466	E4	8.10
24)	Chlordane-2	4.691	4.704	4.603	4.923	5.472	4.879	E4	7.21
25)	Chlordane-3	1.929	1.977	1.913	1.934	2.122	1.975	E5	4.33
26)	Chlordane-4	2.304	2.355	2.294	2.359	2.669	2.396	E5	6.47
27)	Chlordane-5	4.210	4.196	4.074	4.210	4.252	4.189	E4	1.61
28) SA	Decachlorobiphenyl	1.152	1.098	1.092	1.219	1.449	1.202	E6	12.26

(#) = Out of Range