

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093019\
 Data File : PD055068.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2019 22:56
 Operator : SG\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 01 07:43:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX093019.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 01 07:42:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	849.998	858.394 E3	-1.0	102	0.00
2	Toxaphene-1	5.884	6.139 E3	-4.3	103	0.00
3	Toxaphene-2	14.891	15.894 E3	-6.7	104	0.00
4	Toxaphene-3	15.303	16.098 E3	-5.2	102	0.00
5	Toxaphene-4	11.479	12.159 E3	-5.9	102	0.00
6	Toxaphene-5	19.376	19.774 E3	-2.1	101	0.00
7 SA	Decachlorobiphenyl	929.073	916.630 E3	1.3	101	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	1.233	1.260 E6	-2.2	103	0.00
2	Toxaphene-1	6.898	6.697 E3	2.9	93	0.00
3	Toxaphene-2	8.486	8.871 E3	-4.5	101	0.00
4	Toxaphene-3	35.952	36.453 E3	-1.4	102	0.00
5	Toxaphene-4	21.976	22.696 E3	-3.3	103	0.00
6	Toxaphene-5	22.244	21.424 E3	3.7	100	0.00
7 SA	Decachlorobiphenyl	1.303	1.295 E6	0.6	102	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0