

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : DTX093019.M
 Title : GC Extractables
 Last Update : Tue Oct 01 07:22:24 2019
 Response Via : Initial Calibration

Calibration Files

500 =PD055063.D 1000 =PD055061.D 750 =PD055062.D
 250 =PD055064.D 100 =PD055065.D

Compound		500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	8.406	8.386	8.226	8.490	8.991	8.500	E5	3.42
2)	Toxaphene-1	5.969	5.717	5.835	5.993	5.908	5.884	E3	1.90
3)	Toxaphene-2	1.534	1.511	1.514	1.527	1.359	1.489	E4	4.92
4)	Toxaphene-3	1.576	1.535	1.570	1.604	1.366	1.530	E4	6.21
5)	Toxaphene-4	1.191	1.215	1.254	1.134	0.946	1.148	E4	10.55
6)	Toxaphene-5	1.953	1.870	1.903	1.961	1.998	1.937	E4	2.61
7) SA	Decachlorobiphenyl	0.904	0.872	0.877	0.951	1.041	0.929	E6	7.54

Signal #2 Calibration Files

500 =PD055063.D 1000 =PD055061.D 750 =PD055062.D
 250 =PD055064.D 100 =PD055065.D

Compound		500	1000	750	250	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.229	1.230	1.200	1.224	1.281	1.233	E6	2.40
2)	Toxaphene-1	7.227	6.521	6.613	6.931	7.197	6.898	E3	4.71
3)	Toxaphene-2	8.807	8.444	8.273	8.712	8.192	8.486	E3	3.16
4)	Toxaphene-3	3.589	3.491	3.436	3.674	3.786	3.595	E4	3.90
5)	Toxaphene-4	2.212	2.122	2.129	2.245	2.279	2.198	E4	3.18
6)	Toxaphene-5	2.137	2.030	2.066	2.269	2.515	2.203	E4	8.93
7) SA	Decachlorobiphenyl	1.267	1.208	1.218	1.341	1.483	1.303	E6	8.70

(#) = Out of Range