

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093020\
 Data File : PD059440.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2020 23:32
 Operator : DD\AJ
 Sample : INDA337
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3037

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 13:51:24 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD093020CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 30 06:37:17 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.512	4.026	27257608	37655735	19.782	19.793
27)	SA Decachlor...	8.322	9.114	38934984	54510565	39.095	40.953
Target Compounds							
2)	A alpha-BHC	3.954	4.417	39910932	55752598	19.451	19.458
3)	MA gamma-BHC...	4.241	4.702	37436011	50294503	19.261	19.757
4)	MA Heptachlor	4.551	5.215	36814395	47946521	19.622	19.453
9)	A Endosulfan I	5.593	6.266	28528670	37740846	20.244	19.942
13)	MA Dieldrin	5.840	6.524	60078558	82422676	39.519	39.895
14)	MA Endrin	6.099	6.742	47330910	63040113	37.674	38.045
16)	A 4,4'-DDD	6.224	6.861	47220129	62937586	38.672	40.213
17)	MA 4,4'-DDT	6.469	7.161	49865132	63366378	39.029	39.031
20)	A Methoxychlor	7.019	7.617	117.7E6	162.5E6	194.910	196.970

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093020\
Data File : PD059440.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2020 23:32
Operator : DD\AJ
Sample : INDA337
Misc :
ALS Vial : 26 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 13:51:24 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD093020CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 30 06:37:17 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

