

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093022\
 Data File : PD072129.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2022 11:45
 Operator : AR\AJ
 Sample : PCHLORICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 11:57:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD093022.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 30 11:50:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.802	3.562	70834148	746.1E6	24.225	26.483
28)	SA Decachlor...	7.990	9.085	57665827	239.4E6	25.993	27.010
Target Compounds							
23)	Chlordane-1	3.814	4.725	24081639	252.9E6	249.595	276.978
24)	Chlordane-2	4.396	5.257	26572421	184.6E6	265.940	287.002
25)	Chlordane-3	5.032	5.969	78568482	565.6E6	241.900	267.716
26)	Chlordane-4	5.096	6.047	72243643	660.0E6	242.671	268.231
27)	Chlordane-5	5.999	6.908	27028544	127.9E6	247.511	263.777



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093022\
Data File : PD072129.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2022 11:45
Operator : AR\AJ
Sample : PCHLORICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 11:57:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD093022.M
Quant Title : GC Extractables
QLast Update : Fri Sep 30 11:50:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 2.802 min
Delta R.T.: 0.000 min
Response: 70834148
Conc: 24.23 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.562 min
Delta R.T.: 0.000 min
Response: 746063819
Conc: 26.48 ng/ml




#23 Chlordane-1

R.T.: 3.814 min
Delta R.T.: 0.000 min
Response: 24081639
Conc: 249.60 ng/ml



#23 Chlordane-1


R.T.: 4.725 min
Delta R.T.: 0.000 min
Response: 252895340
Conc: 276.98 ng/ml



#24 Chlordane-2


R.T.: 4.396 min
Delta R.T.: 0.000 min
Response: 26572421
Conc: 265.94 ng/ml

Instrument :
ECD_D
ClientSampleId :




#24 Chlordane-2

R.T.: 5.257 min
Delta R.T.: 0.000 min
Response: 184554863
Conc: 287.00 ng/ml




#25 Chlordane-3

R.T.: 5.032 min
Delta R.T.: 0.000 min
Response: 78568482
Conc: 241.90 ng/ml



#25 Chlordane-3


R.T.: 5.969 min
Delta R.T.: 0.000 min
Response: 565592894
Conc: 267.72 ng/ml



#26 Chlordane-4


R.T.: 5.096 min
Delta R.T.: 0.000 min
Response: 72243643
Conc: 242.67 ng/ml

Instrument :
ECD_D
ClientSampleId :




#26 Chlordane-4

R.T.: 6.047 min
Delta R.T.: 0.000 min
Response: 659985318
Conc: 268.23 ng/ml




#27 Chlordane-5

R.T.: 5.999 min
Delta R.T.: 0.000 min
Response: 27028544
Conc: 247.51 ng/ml



#27 Chlordane-5


R.T.: 6.908 min
Delta R.T.: 0.000 min
Response: 127882055
Conc: 263.78 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.990 min
Delta R.T.: 0.000 min
Response: 57665827
Conc: 25.99 ng/ml

Instrument :
ECD_D
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.085 min
Delta R.T.: 0.000 min
Response: 239360422
Conc: 27.01 ng/ml