

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093022\
 Data File : PD072121.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Sep 2022 09:56
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 10:37:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD093022.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 30 10:36:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.803	3.562	250.4E6	2864.3E6	102.673	98.424
28)	SA Decachlor...	7.990	9.084	218.1E6	873.7E6	98.158	94.706
Target Compounds							
2)	A alpha-BHC	3.312	4.028	415.4E6	4071.2E6	105.656	99.134
3)	MA gamma-BHC...	3.645	4.363	374.3E6	3250.1E6	104.847	97.933
4)	MA Heptachlor	3.991	4.940	380.5E6	2294.1E6	103.059	94.922
5)	MB Aldrin	4.274	5.283	371.6E6	1899.4E6	103.262	94.943
6)	B beta-BHC	3.943	4.570	147.8E6	1240.8E6	99.915	94.975
7)	B delta-BHC	4.176	4.818	370.7E6	2467.5E6	105.032	100.465
8)	B Heptachlo...	4.780	5.714	324.2E6	1518.2E6	102.255	96.414
9)	A Endosulfan I	5.153	6.102	301.0E6	1398.5E6	101.986	95.720
10)	B gamma-Chl...	5.033	5.969	336.1E6	1529.7E6	103.760	96.537
11)	B alpha-Chl...	5.098	6.046	327.2E6	1478.1E6	102.780	95.944
12)	B 4,4'-DDE	5.288	6.219	310.1E6	1423.1E6	104.203	96.934
13)	MA Dieldrin	5.421	6.379	324.5E6	1469.0E6	103.536	96.292
14)	MA Endrin	5.698	6.612	279.5E6	1202.8E6	102.447	96.349
15)	B Endosulfa...	5.993	6.837	251.3E6	1138.8E6	101.640	95.688
16)	A 4,4'-DDD	5.846	6.743	254.4E6	1158.9E6	102.406	96.837
17)	MA 4,4'-DDT	6.099	7.054	262.0E6	1183.8E6	104.512	97.824
18)	B Endrin al...	6.173	6.968	196.1E6	904.3E6	99.651	95.634
19)	B Endosulfa...	6.397	7.201	248.0E6	1086.6E6	100.317	96.495
20)	A Methoxychlor	6.678	7.530	134.4E6	576.5E6	97.732	94.863
21)	B Endrin ke...	6.907	7.693	290.9E6	1200.6E6	100.846	96.077
22)	Mirex	7.093	8.154	212.0E6	802.7E6	98.098	95.127

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093022\
Data File : PD072121.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Sep 2022 09:56
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 10:37:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD093022.M
Quant Title : GC Extractables
QLast Update : Fri Sep 30 10:36:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

