

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093024\
 Data File : PD085814.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Oct 2024 01:17
 Operator : AR\AJ
 Sample : P4020-04DL 10X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDCK6DL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/02/2024
 Supervised By : Ankita Jodhani 10/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 21:30:38 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092724CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 27 15:18:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.439	2.888	2610474	80328670	2.492m	18.972m#
2) SA Decachlor...	8.884	8.061	3938391	2420025	2.672m	0.626m#
Target Compounds						
2) A alpha-BHC	3.886	3.386	158.3E6	407.7E6	87.002	63.123 #
5) MB Aldrin	5.121	4.344	7874996	20631017	4.683m	3.466m#
7) B delta-BHC	4.644	4.250	237.9E6	428.5E6	134.436	71.878m#
10) B trans-Chl...	5.801	5.113	3048024	9949750	2.042m	1.838m
11) B cis-Chlor...	5.887	5.183	3842461	16086171	2.464m	2.911m
13) MA Dieldrin	6.219	5.525	6155962	31734651	3.771m	5.617m#
16) A 4,4'-DDD	6.576	5.917	105.0E6	278.0E6	95.342m	65.387m#
17) MA 4,4'-DDT	6.890	6.171	17103456	58176204	15.389m	13.310m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD093024\
Data File : PD085814.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Oct 2024 01:17
Operator : AR\AJ
Sample : P4020-04DL 10X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 21:30:38 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Sep 27 15:18:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Miza 10/02/2024
Supervised By: Ankita Jodhani 10/02/2024

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
DDCK6DL