

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100119\
 Data File : PD055131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2019 03:59
 Operator : SG\AJ
 Sample : PIBLK46
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PIBLK46

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 02 07:36:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092319CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 24 07:21:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.292	3.963	17195072	24014838	22.602	23.402
27)	SA Decachlor...	7.970	8.998	33009151	41784676	39.176	35.865
Target Compounds							
8)	B Heptachlo...	4.960	0.000	129699	0	0.141	N.D. #
12)	B 4,4'-DDE	5.436	0.000	633990	0	0.707	N.D. #
13)	MA Dieldrin	5.520	0.000	274912	0	0.279	N.D. #
22)	Toxaphene-1	5.520	0.000	274912	0	49.142	N.D. #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100119\
Data File : PD055131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2019 03:59
Operator : SG\AJ
Sample : PIBLK46
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 02 07:36:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD092319CLP.M
Quant Title : GC Extractables
QLast Update : Tue Sep 24 07:21:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




Instrument :
ECD_D
ClientSampleId :
PIBLK46



#1 Tetrachloro-m-xylene


R.T.: 3.292 min
Delta R.T.: -0.003 min
Response: 17195072
Conc: 22.60 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK46




#1 Tetrachloro-m-xylene

R.T.: 3.963 min
Delta R.T.: -0.004 min
Response: 24014838
Conc: 23.40 ng/ml




#8 Heptachlor epoxide

R.T.: 4.960 min
Delta R.T.: -0.001 min
Response: 129699
Conc: 0.14 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.831 min
Response: 0
Conc: N.D.



#12 4,4'-DDE


R.T.: 5.436 min
Delta R.T.: 0.026 min
Response: 633990
Conc: 0.71 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK46




#12 4,4'-DDE

R.T.: 0.000 min
Exp R.T. : 6.293 min
Response: 0
Conc: N.D.




#13 Dieldrin

R.T.: 5.520 min
Delta R.T.: -0.012 min
Response: 274912
Conc: 0.28 ng/ml



#13 Dieldrin


R.T.: 0.000 min
Exp R.T. : 6.441 min
Response: 0
Conc: N.D.



#22 Toxaphene-1


R.T.: 5.520 min
Delta R.T.: 0.024 min
Response: 274912
Conc: 49.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK46




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T.: 6.342 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 7.970 min
Delta R.T.: -0.007 min
Response: 33009151
Conc: 39.18 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.998 min
Delta R.T.: -0.006 min
Response: 41784676
Conc: 35.86 ng/ml