

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100322\
 Data File : PD072165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Oct 2022 18:14
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 04 04:21:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100322.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 04 04:21:28 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.803	3.562	176.3E6	2204.1E6	74.214	74.122
28)	SA Decachlor...	7.988	9.084	165.7E6	609.4E6	74.204	74.237
Target Compounds							
2)	A alpha-BHC	3.311	4.028	287.0E6	3086.0E6	74.331	74.073
3)	MA gamma-BHC...	3.644	4.364	257.8E6	2476.1E6	74.366	74.046
4)	MA Heptachlor	3.990	4.940	260.1E6	1694.4E6	74.369	74.113
5)	MB Aldrin	4.273	5.283	255.5E6	1428.3E6	74.547	74.057
6)	B beta-BHC	3.943	4.571	103.3E6	947.7E6	74.552	73.875
7)	B delta-BHC	4.176	4.818	256.0E6	1689.8E6	74.525	73.592
8)	B Heptachlo...	4.779	5.714	223.2E6	1052.2E6	74.366	73.891
9)	A Endosulfan I	5.152	6.101	208.6E6	1021.5E6	74.317	74.307
10)	B gamma-Chl...	5.032	5.969	227.3E6	1099.9E6	74.261	74.134
11)	B alpha-Chl...	5.096	6.046	223.5E6	1067.3E6	74.270	74.061
12)	B 4,4'-DDE	5.287	6.219	210.9E6	1015.2E6	74.383	74.107
13)	MA Dieldrin	5.419	6.378	223.6E6	1060.3E6	74.422	74.157
14)	MA Endrin	5.696	6.611	192.2E6	852.4E6	74.672	74.304
15)	B Endosulfa...	5.991	6.837	177.4E6	812.5E6	74.380	74.094
16)	A 4,4'-DDD	5.844	6.743	174.8E6	808.4E6	74.298	73.899
17)	MA 4,4'-DDT	6.098	7.054	179.2E6	826.8E6	74.814	74.306
18)	B Endrin al...	6.172	6.968	142.0E6	646.9E6	74.250	73.954
19)	B Endosulfa...	6.396	7.201	177.0E6	760.8E6	74.495	74.375
20)	A Methoxychlor	6.676	7.531	96899465	398.2E6	74.564	74.155
21)	B Endrin ke...	6.905	7.692	215.5E6	852.7E6	74.671	74.214
22)	Mirex	7.090	8.153	153.8E6	565.6E6	74.186	74.305

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100322\
Data File : PD072165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Oct 2022 18:14
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 04 04:21:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100322.M
Quant Title : GC Extractables
QLast Update : Tue Oct 04 04:21:28 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

