

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100324\
 Data File : PD085894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Oct 2024 15:58
 Operator : AR\AJ
 Sample : P4022-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDCA1

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/03/2024
 Supervised By :Ankita Jodhani 10/03/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 01:37:57 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100124CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 01 08:04:20 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	2.887	136.1E6	203.5E6	118.982m	47.997m#
27)	SA Decachlor...	8.887	8.058	35479667	81821019	19.509	22.259
Target Compounds							
2)	A alpha-BHC	3.883	3.383	270.8E6	234.2E6	130.061m	37.122m#
7)	B delta-BHC	4.640	4.249	49607655	558.7E6	25.818m	94.136 #
8)	B Heptachlo...	5.566	4.852	37876342	525.9E6	19.839m	96.872m#
11)	B cis-Chlor...	5.889	5.164	24016347	104.2E6	12.447m	18.981 #
13)	MA Dieldrin	6.213	5.486	57249518	342.0E6	28.573m	61.700m#
16)	A 4,4'-DDD	6.597	5.911	36363482	74942327	29.158m	19.157m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100324\
Data File : PD085894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Oct 2024 15:58
Operator : AR\AJ
Sample : P4022-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 01:37:57 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100124CLP.M
Quant Title : GC Extractables
QLast Update : Tue Oct 01 08:04:20 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Mirza 10/03/2024
Supervised By: Ankita Jodhani 10/03/2024

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampled :
DDCA1