

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD100418CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 05 01:49:58 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049571.D 1 =PD049569.D 2 =PD049570.D
 4 =PD049572.D 5 =PD049573.D

Compound		3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.209	1.291	1.269	1.238	1.263	1.254	E6	2.52
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	6.641	6.989	7.432	6.965	6.769	6.959	E3	4.32
23)	Toxaphene-2	8.110	9.049	8.701	8.569	9.007	8.687	E3	4.39
24)	Toxaphene-3	2.233	2.340	2.272	2.400	2.588	2.366	E4	5.90
25)	Toxaphene-4	2.278	2.416	2.318	2.470	2.775	2.451	E4	8.01
26)	Toxaphene-5	2.726	2.861	2.773	2.867	3.063	2.858	E4	4.52
27) SA	Decachlorobiphenyl	1.116	1.265	1.164	1.048	1.039	1.127	E6	8.26

Signal #2 Calibration Files

3 =PD049571.D 1 =PD049569.D 2 =PD049570.D
 4 =PD049572.D 5 =PD049573.D

Compound		3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.051	1.123	1.084	1.024	0.995	1.056	E7	4.75
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	4.630	4.757	5.524	4.261	4.418	4.718	E4	10.37
23)	Toxaphene-2	5.577	5.495	5.753	5.608	5.311	5.549	E4	2.93

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD100418CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 05 01:49:58 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049571.D 1 =PD049569.D 2 =PD049570.D
 4 =PD049572.D 5 =PD049573.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3	2.724	4.383	2.757	2.636	2.693	3.039 E5	24.77
25)	Toxaphene-4	5.273	5.416	5.164	5.169	5.316	5.267 E4	2.01
26)	Toxaphene-5	9.508	9.843	9.470	9.206	9.213	9.448 E4	2.77
27) SA	Decachlorobiphenyl	8.801	9.503	8.982	8.029	7.934	8.650 E6	7.67

(#) = Out of Range