

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD100418CLP.M
 Title : GC Extractables
 Last Update : Fri Oct 05 02:46:31 2018
 Response Via : Initial Calibration

Calibration Files

3 =PD049578.D 1 =PD049574.D 2 =PD049576.D
 4 =PD049580.D 5 =PD049582.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.177	1.220	1.219	1.173	1.167	1.191 E6	2.19
2) A	alpha-BHC	1.648	1.579	1.618	1.741	1.813	1.680 E6	5.67
3) MA	gamma-BHC (Lindane)	1.580	1.606	1.594	1.628	1.670	1.615 E6	2.18
4) MA	Heptachlor	1.567	1.597	1.593	1.614	1.645	1.603 E6	1.79
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I	1.303	1.339	1.341	1.317	1.326	1.325 E6	1.19
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin	1.419	1.380	1.409	1.485	1.539	1.446 E6	4.45
14) MA	Endrin	1.172	1.155	1.176	1.215	1.256	1.195 E6	3.42
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD	1.044	1.015	1.040	1.088	1.128	1.063 E6	4.21
17) MA	4,4'-DDT	1.068	1.031	1.063	1.121	1.170	1.090 E6	5.05
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor	6.176	6.211	6.260	6.260	6.321	6.246 E5	0.88
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	1.065	1.191	1.132	1.011	0.985	1.077 E6	7.90

Signal #2 Calibration Files

3 =PD049578.D 1 =PD049574.D 2 =PD049576.D
 4 =PD049580.D 5 =PD049582.D

	Compound	3	1	2	4	5	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.014	1.067	1.026	1.002	0.957	1.013 E7	3.94
2) A	alpha-BHC	1.438	1.410	1.448	1.441	1.407	1.429 E7	1.32
3) MA	gamma-BHC (Lindane)	1.309	1.328	1.327	1.298	1.255	1.303 E7	2.28
4) MA	Heptachlor	1.266	1.313	1.307	1.248	1.207	1.268 E7	3.46
5) MB	Aldrin						0.000	-1.00
6) B	beta-BHC						0.000	-1.00
7) B	delta-BHC						0.000	-1.00
8) B	Heptachlor epoxide						0.000	-1.00
9) A	Endosulfan I	0.968	1.023	1.014	0.953	0.912	0.974 E7	4.69
10) B	trans-Chlordane						0.000	-1.00
11) B	cis-Chlordane						0.000	-1.00
12) B	4,4'-DDE						0.000	-1.00
13) MA	Dieldrin	1.021	1.133	1.021	1.023	0.990	1.038 E7	5.31
14) MA	Endrin	7.130	7.182	7.106	6.912	6.842	7.034 E6	2.11
15) B	Endosulfan II						0.000	-1.00
16) A	4,4'-DDD	7.634	7.977	7.779	7.671	7.625	7.737 E6	1.91
17) MA	4,4'-DDT	6.937	6.968	6.988	7.130	7.207	7.046 E6	1.66
18) B	Endrin aldehyde						0.000	-1.00
19) B	Endosulfan Sulfate						0.000	-1.00
20) A	Methoxychlor	3.353	3.386	3.357	3.358	3.387	3.368 E6	0.49
21) B	Endrin ketone						0.000	-1.00
22)	Toxaphene-1						0.000	-1.00
23)	Toxaphene-2						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD100418CLP.M
Title : GC Extractables
Last Update : Fri Oct 05 02:46:31 2018
Response Via : Initial Calibration

Calibration Files

3 =PD049578.D 1 =PD049574.D 2 =PD049576.D
4 =PD049580.D 5 =PD049582.D

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	8.370	8.593	8.643	8.030	7.748	8.277 E6	4.62

(#) = Out of Range