

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100423\
 Data File : PD078615.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Oct 2023 16:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 21:25:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100323.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 03 01:10:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.540	2.758	146.3E6	103.8E6	50.667	50.817
28)	SA Decachlor...	9.081	7.895	188.9E6	117.4E6	50.932	51.735
Target Compounds							
2)	A alpha-BHC	3.997	3.259	251.0E6	156.9E6	51.685	52.339
3)	MA gamma-BHC...	4.330	3.588	242.2E6	152.2E6	51.783	52.530
4)	MA Heptachlor	4.921	3.927	241.2E6	149.6E6	51.305	50.716
5)	MB Aldrin	5.264	4.206	250.0E6	154.3E6	51.758	52.008
6)	B beta-BHC	4.528	3.885	93485445	59778785	50.334	51.723
7)	B delta-BHC	4.776	4.115	242.9E6	152.9E6	53.918	53.447
8)	B Heptachlo...	5.691	4.708	217.9E6	137.6E6	51.890	50.920
9)	A Endosulfan I	6.079	5.078	200.5E6	124.7E6	51.321	49.607
10)	B gamma-Chl...	5.949	4.959	218.1E6	136.9E6	51.937	50.952
11)	B alpha-Chl...	6.028	5.023	216.7E6	136.7E6	51.445	50.925
12)	B 4,4'-DDE	6.201	5.211	207.5E6	135.4E6	51.845	53.228
13)	MA Dieldrin	6.354	5.343	218.2E6	143.1E6	52.000	51.248
14)	MA Endrin	6.585	5.619	173.6E6	126.6E6	51.700	51.170
15)	B Endosulfa...	6.805	5.913	181.9E6	121.6E6	50.542	51.075
16)	A 4,4'-DDD	6.719	5.766	162.2E6	109.1E6	54.522	55.149
17)	MA 4,4'-DDT	7.035	6.017	156.2E6	106.6E6	52.770	52.363
18)	B Endrin al...	6.936	6.093	133.6E6	93688888	50.027	50.537
19)	B Endosulfa...	7.171	6.316	167.6E6	114.4E6	52.157	51.587
20)	A Methoxychlor	7.511	6.593	86293774	57548872	51.706	52.747
21)	B Endrin ke...	7.657	6.823	191.9E6	139.1E6	53.401	52.985
22)	Mirex	8.131	7.003	139.6E6	102.8E6	52.278	50.438

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100423\
Data File : PD078615.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Oct 2023 16:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 21:25:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100323.M
Quant Title : GC Extractables
QLast Update : Tue Oct 03 01:10:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

