

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100423\
 Data File : PD078628.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Oct 2023 19:56
 Operator : AR\AJ
 Sample : 04685-05MSD
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PP-20230928-B6A-0-7.5MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 10/04/2023
 Supervised By :Ankita Jodhani 10/04/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 03 21:36:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100323.M
 Quant Title : GC Extractables
 QLast Update : Tue Oct 03 01:10:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.540	2.758	48348152	34436572	16.748	16.861
28) SA Decachlor...	9.078	7.894	56151006	32422722	15.141	14.289
Target Compounds						
2) A alpha-BHC	3.997	3.259	240.9E6	149.0E6	49.613	49.719
3) MA gamma-BHC...	4.330	3.588	232.3E6	146.4E6	49.666	50.523
4) MA Heptachlor	4.920	3.927	229.0E6	142.7E6	48.698	48.350
5) MB Aldrin	5.263	4.206	239.3E6	146.1E6	49.533	49.237
6) B beta-BHC	4.528	3.885	91600534	57044881	49.319	49.358
7) B delta-BHC	4.776	4.115	228.6E6	144.7E6	50.748	50.568
8) B Heptachlo...	5.690	4.708	207.0E6	134.1E6	49.275	49.613
9) A Endosulfan I	6.077	5.078	192.4E6	124.5E6	49.239	49.522
10) B gamma-Chl...	5.947	4.958	207.6E6	133.9E6	49.434	49.807
11) B alpha-Chl...	6.026	5.022	207.2E6	132.8E6	49.181	49.456
12) B 4,4'-DDE	6.199	5.210	199.0E6	129.5E6	49.714	50.911
13) MA Dieldrin	6.353	5.342	208.1E6	138.2E6	49.610	49.502
14) MA Endrin	6.584	5.618	172.8E6	123.1E6	51.442	49.760
15) B Endosulfa...	6.803	5.912	170.5E6	116.7E6	47.373	49.031
16) A 4,4'-DDD	6.718	5.765	155.8E6	104.5E6	52.350	52.797
17) MA 4,4'-DDT	7.032	6.015	150.8E6	103.4E6	50.944	50.817
18) B Endrin al...	6.932	6.092	130.9E6	89539256	48.984m	48.298
19) B Endosulfa...	7.169	6.314	158.1E6	109.9E6	49.211	49.535
20) A Methoxychlor	7.509	6.591	81947942	56288715	49.102	51.592
21) B Endrin ke...	7.655	6.821	184.9E6	129.8E6	51.475	49.467
22) Mirex	8.129	7.002	152.3E6	113.6E6	57.037	55.751

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100423\
Data File : PD078628.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Oct 2023 19:56
Operator : AR\AJ
Sample : 04685-05MSD
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PP-20230928-B6A-0-7.5MSD

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 10/04/2023
Supervised By :Ankita Jodhani 10/04/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 03 21:36:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100323.M
Quant Title : GC Extractables
QLast Update : Tue Oct 03 01:10:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

