

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
 Data File : PD072216.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 11:44
 Operator : AR\AJ
 Sample : PCHLORICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 12:30:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 05 12:29:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.744	3.497	152.2E6	1765.4E6	50.000	50.000
28)	SA Decachlor...	7.920	8.986	126.3E6	497.9E6	50.000	50.000
Target Compounds							
23)	Chlordane-1	3.749	4.653	49458545	574.9E6	500.000	500.000
24)	Chlordane-2	4.329	5.185	56534383	449.7E6	500.000	500.000
25)	Chlordane-3	4.964	5.895	160.4E6	1233.2E6	500.000	500.000
26)	Chlordane-4	5.028	5.972	154.2E6	1395.9E6	500.000	500.000
27)	Chlordane-5	5.927	6.831	52941613	270.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
Data File : PD072216.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 11:44
Operator : AR\AJ
Sample : PCHLORICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 12:30:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
Quant Title : GC Extractables
QLast Update : Wed Oct 05 12:29:12 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 2.744 min
Delta R.T.: 0.000 min
Response: 152153946
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.497 min
Delta R.T.: 0.000 min
Response: 1765410788
Conc: 50.00 ng/ml




#23 Chlordane-1

R.T.: 3.749 min
Delta R.T.: 0.000 min
Response: 49458545
Conc: 500.00 ng/ml



#23 Chlordane-1


R.T.: 4.653 min
Delta R.T.: 0.000 min
Response: 574902302
Conc: 500.00 ng/ml



#24 Chlordane-2


R.T.: 4.329 min
Delta R.T.: 0.000 min
Response: 56534383
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#24 Chlordane-2

R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 449655073
Conc: 500.00 ng/ml




#25 Chlordane-3

R.T.: 4.964 min
Delta R.T.: 0.000 min
Response: 160425179
Conc: 500.00 ng/ml



#25 Chlordane-3


R.T.: 5.895 min
Delta R.T.: 0.000 min
Response: 1233193795
Conc: 500.00 ng/ml



#26 Chlordane-4


R.T.: 5.028 min
Delta R.T.: 0.000 min
Response: 154229863
Conc: 500.00 ng/ml

Instrument :
ECD_D
ClientSampleId :




#26 Chlordane-4

R.T.: 5.972 min
Delta R.T.: 0.000 min
Response: 1395913920
Conc: 500.00 ng/ml




#27 Chlordane-5

R.T.: 5.927 min
Delta R.T.: 0.000 min
Response: 52941613
Conc: 500.00 ng/ml



#27 Chlordane-5


R.T.: 6.831 min
Delta R.T.: 0.000 min
Response: 270378431
Conc: 500.00 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.920 min
Delta R.T.: 0.000 min
Response: 126288308
Conc: 50.00 ng/ml

Instrument :
ECD_D
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 8.986 min
Delta R.T.: 0.000 min
Response: 497898329
Conc: 50.00 ng/ml