

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
 Data File : PD072209.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 10:10
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 11:57:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 05 11:57:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.744	3.497	257.6E6	3466.1E6	102.139	97.658
28)	SA Decachlor...	7.920	8.985	241.1E6	938.0E6	96.837	94.809
Target Compounds							
2)	A alpha-BHC	3.249	3.960	434.7E6	4937.1E6	105.583	98.670
3)	MA gamma-BHC...	3.580	4.294	389.0E6	3906.9E6	104.932	97.675
4)	MA Heptachlor	3.924	4.869	394.5E6	2740.8E6	103.641	95.111
5)	MB Aldrin	4.205	5.209	373.3E6	2181.4E6	103.974	94.930
6)	B beta-BHC	3.879	4.502	154.8E6	1523.5E6	99.525	95.100
7)	B delta-BHC	4.110	4.747	393.1E6	2897.4E6	105.119	98.407
8)	B Heptachlo...	4.710	5.640	327.6E6	1719.2E6	102.268	95.052
9)	A Endosulfan I	5.083	6.026	304.1E6	1561.2E6	102.324	94.796
10)	B gamma-Chl...	4.963	5.894	341.8E6	1722.9E6	104.063	96.212
11)	B alpha-Chl...	5.027	5.971	331.2E6	1661.4E6	103.442	96.011
12)	B 4,4'-DDE	5.220	6.146	324.7E6	1652.8E6	105.377	96.497
13)	MA Dieldrin	5.349	6.302	331.9E6	1638.0E6	104.253	96.693
14)	MA Endrin	5.625	6.534	283.0E6	1348.9E6	104.635	96.348
15)	B Endosulfa...	5.921	6.761	263.6E6	1280.1E6	102.444	96.262
16)	A 4,4'-DDD	5.777	6.670	273.4E6	1335.2E6	104.702	96.616
17)	MA 4,4'-DDT	6.030	6.980	283.1E6	1368.3E6	105.309	97.038
18)	B Endrin al...	6.102	6.891	208.7E6	1012.5E6	99.473	95.244
19)	B Endosulfa...	6.326	7.124	267.6E6	1210.3E6	98.978	96.896
20)	A Methoxychlor	6.610	7.457	149.6E6	657.1E6	96.990	95.613
21)	B Endrin ke...	6.834	7.614	320.9E6	1327.6E6	100.384	96.203
22)	Mirex	7.018	8.073	224.8E6	870.3E6	99.386	95.326

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
Data File : PD072209.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 10:10
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 11:57:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
Quant Title : GC Extractables
QLast Update : Wed Oct 05 11:57:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

