

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
 Data File : PD072212.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 10:50
 Operator : AR\AJ
 Sample : PSTDICC025
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC025

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/06/2022
 Supervised By : Ankita Jodhani 10/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 11:58:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 05 11:57:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.744	3.497	61973139	913.6E6	24.571	25.741
28)	SA Decachlor...	7.920	8.985	64044523	256.8E6	25.727	25.960
Target Compounds							
2)	A alpha-BHC	3.249	3.960	96477981	1272.5E6	23.435	25.432
3)	MA gamma-BHC...	3.580	4.294	87757874	1034.1E6	23.675	25.853
4)	MA Heptachlor	3.924	4.868	91694441	769.1E6	24.089	26.690
5)	MB Aldrin	4.205	5.209	86201494	604.5E6	24.008	26.306
6)	B beta-BHC	3.879	4.502	39233508	427.6E6	25.222	26.691
7)	B delta-BHC	4.111	4.747	88001612	748.0E6	23.533	25.405
8)	B Heptachlo...	4.711	5.640	78230480	481.2E6	24.424	26.607
9)	A Endosulfan I	5.082	6.026	73016195	438.0E6	24.565	26.593
10)	B gamma-Chl...	4.963	5.894	79589184	471.5E6	24.235	26.330
11)	B alpha-Chl...	5.027	5.971	78332296	457.4E6	24.465	26.431
12)	B 4,4'-DDE	5.220	6.147	73667147	444.1E6	23.906	25.929
13)	MA Dieldrin	5.349	6.303	76006816	441.1E6	23.873	26.037
14)	MA Endrin	5.626	6.535	66033678	373.2E6	24.419	26.657
15)	B Endosulfa...	5.921	6.761	62700685	341.4E6	24.369	25.672
16)	A 4,4'-DDD	5.777	6.668	63962283	350.7E6	24.496	25.381m
17)	MA 4,4'-DDT	6.030	6.981	63872868	360.6E6	23.761	25.577
18)	B Endrin al...	6.102	6.892	52779634	276.8E6	25.162	26.039
19)	B Endosulfa...	6.327	7.125	66863466	320.5E6	24.734	25.657
20)	A Methoxychlor	6.611	7.458	39076734	177.1E6	25.329	25.774
21)	B Endrin ke...	6.834	7.614	79024016	355.2E6	24.722	25.737
22)	Mirex	7.017	8.074	58095623	237.2E6	25.684	25.984

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
Data File : PD072212.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Oct 2022 10:50
Operator : AR\AJ
Sample : PSTDICC025
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDICC025

Manual Integrations

APPROVED

Reviewed By : Abdul Mirza 10/06/2022
Supervised By : Ankita Jodhani 10/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 05 11:58:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
Quant Title : GC Extractables
QLast Update : Wed Oct 05 11:57:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

