

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
 Data File : PD072229.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 15:39
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/06/2022
 Supervised By : Ankita Jodhani 10/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 16:16:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 05 15:15:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.744	3.497	123.1E6	1649.8E6	49.067	45.510
28)	SA Decachlor...	7.919	8.986	125.2E6	492.0E6	49.129	48.966
Target Compounds							
2)	A alpha-BHC	3.249	3.960	200.6E6	2346.1E6	50.122	46.149
3)	MA gamma-BHC...	3.580	4.294	181.1E6	1896.9E6	50.121	45.758
4)	MA Heptachlor	3.923	4.869	186.9E6	1393.8E6	49.747	46.302
5)	MB Aldrin	4.205	5.209	177.7E6	1133.0E6	50.251	46.946
6)	B beta-BHC	3.878	4.502	76247102	775.7E6	49.306	46.547
7)	B delta-BHC	4.110	4.748	182.9E6	1418.4E6	50.188	47.902
8)	B Heptachlo...	4.711	5.640	158.2E6	893.6E6	49.636	47.061
9)	A Endosulfan I	5.083	6.027	145.7E6	806.9E6	49.059	47.542
10)	B gamma-Chl...	4.963	5.895	161.3E6	875.6E6	49.196	47.235
11)	B alpha-Chl...	5.027	5.972	157.8E6	846.9E6	49.156	47.181
12)	B 4,4'-DDE	5.220	6.147	152.1E6	834.3E6	49.978	46.902
13)	MA Dieldrin	5.349	6.303	155.7E6	832.9E6	49.433	47.809
14)	MA Endrin	5.626	6.536	134.7E6	675.2E6	49.619	47.796
15)	B Endosulfa...	5.920	6.761	127.7E6	648.8E6	50.005m	48.049
16)	A 4,4'-DDD	5.778	6.670	131.1E6	685.6E6	50.210	49.326
17)	MA 4,4'-DDT	6.030	6.980	133.1E6	689.3E6	50.773	48.857
18)	B Endrin al...	6.102	6.891	104.5E6	521.7E6	49.056	48.102
19)	B Endosulfa...	6.326	7.124	133.0E6	615.2E6	49.197	48.584
20)	A Methoxychlor	6.610	7.457	76450904	342.6E6	49.536	49.359
21)	B Endrin ke...	6.833	7.613	160.1E6	686.5E6	50.337	48.971
22)	Mirex	7.017	8.073	113.4E6	455.5E6	48.745	49.078

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100522\
 Data File : PD072229.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Oct 2022 15:39
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 10/06/2022
 Supervised By :Ankita Jodhani 10/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 05 16:16:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100522.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 05 15:15:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

