

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100620\
 Data File : PD059536.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Oct 2020 20:39
 Operator : DD\AJ
 Sample : INDA343
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA343

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 07 08:08:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100220CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Oct 02 04:21:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.511	4.026	26967471	37743085	18.669	19.003
27)	SA Decachlor...	8.316	9.111	39060380	56486584	36.454	37.966
Target Compounds							
2)	A alpha-BHC	3.953	4.416	39733088	56046304	18.713	19.185
3)	MA gamma-BHC...	4.240	4.702	37003144	52816888	18.947	19.902
4)	MA Heptachlor	4.549	5.214	37068896	50334893	18.986	19.806
9)	A Endosulfan I	5.590	6.265	28194426	39708490	18.748	19.469
13)	MA Dieldrin	5.837	6.523	62905800	86828245	38.356	38.950
14)	MA Endrin	6.096	6.741	47234611	65043484	39.985	41.984
16)	A 4,4'-DDD	6.221	6.860	48474679	67799147	37.077	39.163
17)	MA 4,4'-DDT	6.465	7.160	49825263	66638193	37.129	41.517
20)	A Methoxychlor	7.014	7.616	117.6E6	171.9E6	191.622	207.212

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100620\
Data File : PD059536.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Oct 2020 20:39
Operator : DD\AJ
Sample : INDA343
Misc :
ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 07 08:08:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100220CLP.M
Quant Title : GC Extractables
QLast Update : Fri Oct 02 04:21:20 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

