

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD100722\  
 Data File : PD072272.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 07 Oct 2022 09:43  
 Operator : AR\AJ  
 Sample : I.BLK  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 I.BLK

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Oct 07 23:36:15 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD100522.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Oct 05 15:15:26 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-----------------|-------|-------|----------|---------|--------|--------|
| -----                       |                 |       |       |          |         |        |        |
| System Monitoring Compounds |                 |       |       |          |         |        |        |
| 1)                          | SA Tetrachlo... | 2.745 | 3.496 | 49241271 | 736.9E6 | 19.621 | 20.329 |
| 28)                         | SA Decachlor... | 7.921 | 8.986 | 49276266 | 226.3E6 | 19.330 | 22.522 |

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD100722\  
Data File : PD072272.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 07 Oct 2022 09:43  
Operator : AR\AJ  
Sample : I.BLK  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_D  
ClientSampleId :  
I.BLK

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Oct 07 23:36:15 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD100522.M  
Quant Title : GC Extractables  
QLast Update : Wed Oct 05 15:15:26 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 2.745 min  
Delta R.T.: 0.000 min  
Response: 49241271  
Conc: 19.62 ng/ml

Instrument :  
ECD\_D  
ClientSampleId :  
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.496 min  
Delta R.T.: -0.001 min  
Response: 736947378  
Conc: 20.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.921 min  
Delta R.T.: 0.001 min  
Response: 49276266  
Conc: 19.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.986 min  
Delta R.T.: 0.000 min  
Response: 226277530  
Conc: 22.52 ng/ml