

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100920\
 Data File : PD059603.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Oct 2020 15:56
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 09 16:27:44 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100820.M
 Quant Title : GC Extractables
 QLast Update : Thu Oct 08 13:21:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.515	4.025	73411377	107.3E6	50.916	49.345
28)	SA Decachlor...	8.321	9.113	55472540	80642711	49.936	50.048
Target Compounds							
2)	A alpha-BHC	3.958	4.416	119.6E6	170.4E6	52.404	51.191
3)	MA gamma-BHC...	4.245	4.702	110.1E6	158.8E6	52.186	51.242
4)	MA Heptachlor	4.554	5.214	110.7E6	151.1E6	51.626	50.273
5)	MB Aldrin	4.806	5.521	104.9E6	142.0E6	52.811	51.463
6)	B beta-BHC	4.495	4.869	43514938	62826717	51.092	50.416
7)	B delta-BHC	4.704	5.089	104.7E6	148.5E6	52.980	54.019
8)	B Heptachlo...	5.253	5.908	95499912	144.7E6	53.818	56.539
9)	A Endosulfan I	5.596	6.265	85378482	119.0E6	52.808	51.269
10)	B gamma-Chl...	5.483	6.141	93701984	131.3E6	52.804	52.428
11)	B alpha-Chl...	5.542	6.215	92187987	128.5E6	52.514	50.911
12)	B 4,4'-DDE	5.707	6.370	89853123	122.8E6	52.987	51.722
13)	MA Dieldrin	5.841	6.523	92190376	128.1E6	53.519	51.806
14)	MA Endrin	6.101	6.741	74137236	94938794	50.235	45.123
15)	B Endosulfa...	6.375	6.949	78208729	109.4E6	54.045	50.968
16)	A 4,4'-DDD	6.226	6.860	75982501	108.8E6	54.379	53.696
17)	MA 4,4'-DDT	6.470	7.159	75856765	100.1E6	49.771	51.082
18)	B Endrin al...	6.544	7.073	63101539	88667639	51.984	50.308
19)	B Endosulfa...	6.759	7.296	69422993	99510659	53.920	49.909
20)	A Methoxychlor	7.020	7.616	38759098	48881656	51.666	46.574
21)	B Endrin ke...	7.254	7.769	82790594	107.5E6	55.507	50.284
22)	Mirex	7.449	8.225	56424726	67897144	53.129	48.598

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD100920\
Data File : PD059603.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Oct 2020 15:56
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 09 16:27:44 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100820.M
Quant Title : GC Extractables
QLast Update : Thu Oct 08 13:21:41 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

