

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101023\
 Data File : PD078789.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Oct 2023 12:39
 Operator : AR\AJ
 Sample : INDA001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3130

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 11 02:00:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100923CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 11 01:02:10 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.571	2.808	28311911	32996589	18.128	16.384
27)	SA Decachlor...	9.122	7.963	66626072	54112477	36.294	37.212
Target Compounds							
2)	A alpha-BHC	4.028	3.312	47793592	48860451	18.225	16.358
3)	MA gamma-BHC...	4.362	3.642	45574260	46256093	18.229	17.075
4)	MA Heptachlor	4.955	3.986	48008208	43726134	17.847	17.583
9)	A Endosulfan I	6.112	5.141	39433129	35366934	18.406	17.970
13)	MA Dieldrin	6.387	5.406	82923377	76027940	36.574	35.089
14)	MA Endrin	6.618	5.681	66255152	64261346	36.253	35.876
16)	A 4,4'-DDD	6.749	5.828	61514418	56763343	36.581	36.392
17)	MA 4,4'-DDT	7.064	6.080	64628923	55242872	36.470	37.573
20)	A Methoxychlor	7.538	6.656	171.6E6	147.0E6	182.756	188.001

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD101023\
Data File : PD078789.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Oct 2023 12:39
Operator : AR\AJ
Sample : INDA001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 11 02:00:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD100923CLP.M
Quant Title : GC Extractables
QLast Update : Wed Oct 11 01:02:10 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

